

C-Brace Interim.

The **C-Brace Interim** orthosis is a knee-ankle-foot orthosis used primarily in the early stages of rehabilitation and can be used by suitable users for up to 6 months.

Studies show that the use of orthoses during the rehabilitation phase can improve activities of daily living for patients with spinal cord injury. The **C-Brace Interim** orthosis extends the range of exercises available during the rehabilitation process.

- Starting point: Standing ability. The **C-Brace Interim** orthosis can be used as soon as the user is able to stand, or to restore the user's ability to stand.
- Time-saving fitting: Suitable patients can be fitted with the **C-Brace Interim** orthosis within an average of 4 hours (patient fitting time only).
- Flexible downgrade options: The **C-Brace Interim** orthosis offers flexibility in the rehabilitation process. Thanks to its modular design, it can be converted into an ankle-foot orthosis.
- Winning combination: The combination of the interim orthosis and the **C-Brace** orthotic knee joint offers the greatest possible adaptability to the course of therapy.
- Individual flexibility: The use of the **C-Brace** joint unit offers the greatest flexibility to meet individual or changing patient needs.
- Reduced risk of falling: The use of the **C-Brace** joint unit significantly reduces the risk of falling.
- The **C-Brace Interim** orthosis is suitable for unilateral and bilateral use.
- Maximum body weight of the patient: 110 kg
- Body height of the patient: 155 cm to 190 cm



C-Brace Interim.

The **C-Brace Interim** orthosis consists of the following parts:

- Interim AFO (17OK200=*)
- Interim Knee Section (17OK201=*)
- Interim Thigh Shell (17OK202=*)

Interim AFO 17OK200=*

Article Number	Side	Size	Ankle floor measurement	Shoe Size (EU)
17OK200=L-S	Left	S	70 mm	36 – 41
17OK200=R-S	Right	S	70 mm	36 – 41
17OK200=L-L	Left	L	85 mm	39 – 46
17OK200=R-L	Right	L	85 mm	39 – 46

Note: Interim AFO 17OK200=* can be used individually, depending on the patient's fitness and after evaluation by the responsible CPO/CO.

Interim Knee Section 17OK201=*

Article Number	Side	Angle
17OK201=N-0	Neutral	0°
17OK201=N-3	Neutral	3°
17OK201=N-6	Neutral	6°
17OK201=N-9	Neutral	9°

Note: The most favorable variant of the Interim Knee Section 17OK201=* for the patient can be determined using the template set 17OK203=N-2.

Interim Thigh Shell 17OK202=*

Article Number	Side	Size	Circumference (measured 15 cm from center of patella)
17OK202=L-S	Left	S	30 cm – 45 cm
17OK202=R-S	Right	S	30 cm – 45 cm
17OK202=L-M	Left	M	40 cm – 55 cm
17OK202=R-M	Right	M	40 cm – 55 cm
17OK202=L-L	Left	L	50 cm – 65 cm
17OK202=R-L	Right	L	50 cm – 65 cm

Not included in scope:

Article Number	Set	Description
17K01=L-OB 17K01=R-OB	no	C-Brace Loaner Joint provided by Ottobock. Please follow the instructions on order form 647F935.
17OK203=N-1	Tool Set	1x clamping lever, 1x retaining screw, 1x saw guide, 1x clamp adapter
17OK203=N-2	Template Set	4 templates to determine the best-suited Interim Knee Section 17OK201=*

