DEMONSTRATED PROTOCOL

Methanol Fixation of Cells for Single Cell RNA Sequencing

Overview

This protocol outlines methanol fixation and rehydration of single cell suspensions for use with 10x Genomics Single Cell protocols. The protocol was demonstrated with Jurkat T lymphocytes, embryonic brain cells, and human peripheral blood mononuclear cells (PMBCs). Additional optimization may be required when working with other cell types (e.g. media type, resuspension buffer, centrifugation speed, and time).

Preparation of single cell suspensions directly from solid tissues or cryopreserved samples may also require additional optimization during dissociation and/or cell handling, which is not covered here.

Additional Guidance

Consult Demonstrated Protocol Cell Preparation Guide (Document CG00053) for Tips & Best Practices on handling cells. Input single cell suspensions for this protocol were prepared, washed, and counted as described in:

- Demonstrated Protocol Single Cell Suspensions from Cultured Cell Lines for Single Cell RNA Sequencing (Document CG00054)
- Demonstrated Protocol Fresh Frozen Human Peripheral Blood Mononuclear Cells for Single Cell RNA Sequencing (Document CG00039)
- Demonstrated Protocol Dissociation of Mouse Embryonic Neural Tissue for Single Cell RNA Sequencing (Document CG00055)

Cells carry potentially hazardous pathogens. Follow material supplier recommendations and local laboratory procedures and regulations for the safe handling, storage, and disposal of biological materials.

Cell Sourcing

Cell Type Used/Species	Supplier	Part Number
Peripheral Blood T Lymphocytes Jurkat, Clone E6-1/Human	ATCC	ATCC TIB-152
Combined Cortex, Hippocampus, and Ventricular Zone/E18 Mouse	BrainBits	C57EHCV
Peripheral Blood Mononuclear Cells (PBMC)/Human	AllCells	PB001

Preparation – Buffers

Buffers Maintain at 4°C	Composition		
3X SSC Buffer	3X SSC in Nuclease-free Water		
Wash-Resuspension Buffer	0.04% BSA + 1mM DTT + 0.2 U/µl RNase Inhibitor in 3X SSC Buffer		
Additional Buffers/Reagents			
1X DPBS (maintain at 4°C)			
100% Methanol (maintain at -20°C)			

Specific Reagents & Consumables

Vendor	Item	Part Number
Thermo Fisher Scientific	Dulbecco's Phosphate-Buffered Saline (DPBS), No Calcium, No Magnesium	14190144
	Trypan Blue Stain (0.4%)	T10282
	UltraPure Bovine Serum Albumin (BSA, 50 mg/ml)	AM2616
	Countess II FL Automated Cell Counter	AMQAF1000
	Countess II FL Automated Cell Counting Chamber Slides	C10228
Millipore Sigma	Methanol (for HPLC, ≥99.9%)	34860-100ML
	Protector RNAse Inhibitor (40 U/ μ L)	3335399001
	SSC Buffer 20X Concentrate	S6639-1L
	DL-Dithiothreitol Solution BioUltra, for molecular biology	43816
Bel-Art	Flowmi Cell Strainer, 40 µm	H13680-0040
Eppendorf	DNA LoBind Tubes, 2.0 ml	022431048



Protocol Overview





Protocol

1. Methanol Fixation

This protocol was demonstrated using 1-2 x 10⁶ cells.

- a. Harvest or thaw the cells in appropriate culture medium as per manufacturer's instructions. Transfer cells to a 2-ml microcentrifuge tube.
- b. Centrifuge at 300 rcf for 5 min at 4°C.
- c. Remove the supernatant without disrupting the cell pellet.
- d. Using a wide-bore pipette tip, add 1 ml chilled 1X DPBS and gently pipette mix 10x or until cells are resuspended. Each tube should have 1-2 x10⁶ cells.
- e. Centrifuge at 300 rcf for 5 min at 4°C.
- f. Repeat c-e.
- g. Remove the supernatant without disrupting the cell pellet.
- h. Using a wide-bore pipette tip, add 200 µl chilled 1X DPBS/1 x 10⁶ cells and gently pipette mix 10x or until cells are resuspended.
- i. Add 800 µl chilled 100% methanol/1 x 10⁶ cells. To avoid clumping of cells, add methanol drop by drop while gently stirring the cell suspension with the pipette tip in the microcentrifuge tube.

Scale up the volumes of 1X DPBS and methanol if using $>1 \times 10^6$ cells.

- j. Incubate for 30 min at -20°C.
- k. Store fixed cells at -20°C or -80°C for up to 6 weeks or immediately proceed to step 2.

2. Rehydration

- a. Place the microcentrifuge tube containing the methanolfixed cells on ice to equilibrate to 4°C for 5 min.
- b. Centrifuge fixed cells at 1,000 rcf for 5 min at 4°C.
- c. Remove the supernatant without disrupting the cell pellet.
- d. Based on starting cell concentration and assuming ~50% cell loss, add an appropriate volume of Wash-Resuspension Buffer to obtain a concentration of 700-1,200 cells/µl. Gently pipette mix using a regular-bore pipette tip until a single cell suspension is achieved.

If cell debris and large clumps are present, pass the sample through a 40 μm Flowmi Cell Strainer.

- e. Determine cell concentration using a Countess II Automated Cell Counter or a hemocytometer. The targeted final concentration is 700-1,200 cells/µl.
- f. Proceed **immediately** with the 10x Genomics Single Cell protocols. Delay in proceeding may result in RNA loss.

Results

When starting with 2 x 10⁶ total cells, the recovery rate of cells after methanol fixation, storage, and rehydration was approximately 90%. After methanol fixation and rehydration, the cells appeared well singulated with no cellular debris under the microscope. Trypan blue staining of the methanol-fixed and rehydrated cells resulted in all cells staining positively, indicating that the cells have been effectively fixed and permeabilized.

Troubleshooting

Problem	Possible Solution
High fraction of viable cells post methanol fixation	Incrementally increase fixation time and monitor efficacy using microscopy
	Gently pipette mix cell pellet using a 200-µl pipette (set to 100 µl) and a wide-bore pipette tip until no cell clumps are visible in suspension
High fraction of visible debris post methanol fixation	Filter fixed cell suspension with the appropriate strainer Use flow cytometry to sort sample
<700 fixed cells/µl after final resuspension	Concentrate fixed cell suspension to
	achieve target concentration of 700-1,200 fixed cells/µl
	After centrifugation, aspirate supernatant from surface of the tube that is opposite to the surface where the cell pellet forms

References

J. Chen, F. Cheung, R. Shi, H. Zhou, W. Lu, CHI Consortium, PBMC fixation and processing for Chromium Single-Cell RNA sequencing. *J. Transl. Med.* 16, 198 (2018).

© 2019 10X Genomics, Inc. (10x Genomics). All rights reserved. Duplication and/or reproduction of all or any portion of this document without the express written consent of 10x Genomics, is strictly forbidden. Nothing contained herein shall constitute any warranty, express or implied, as to the performance of any products described herein. Any and all warranties applicable to any products are set forth in the applicable terms and conditions of sale accompanying the purchase of such product. 10x Genomics provides no warranty and hereby disclaims any and all warranties as to the use of any third-party products or protocols described herein. The use of products described herein is subject to certain restrictions as set forth in the applicable terms and conditions of sale accompanying the purchase of such product. A non-exhaustive list of 10x Genomics' marks, many of which are registered in the United States and other countries can be viewed at: www.10xgenomics.com/ trademarks. 10x Genomics marks or names. 10x Genomics products may be covered by one or more of the patents as indicated at:www.10xgenomics.com/ patents. The use of products described herein is subject to 10x Genomics products Terms and Conditions of Sale, available at http://www:10xgenomics.com/ patents. The use of products described herein is subject to 10x Genomics and user. All products and services described herein are intended FOR RESEARCH USE ONLY and NOT FOR USE IN DIAGNOSTIC PROCEDURES.

The use of 10x Genomics products in practicing the methods set forth herein has not been validated by 10x Genomics, and such non-validated use is NOT COVERED BY 10X GENOMICS STANDARD WARRANTY, AND 10X GENOMICS HEREBY DISCLAIMS ANY AND ALL WARRANTIES FOR SUCH USE. Nothing in this document should be construed as altering, waiving or amending in any manner 10x Genomics terms and conditions of sale for the Chromium Controller or the Chromium Single Cell Controller, consumables or software, including without limitation such terms and conditions of sale for the Chromium Controller or the Chromium Single Cell Controller, consumables or software, including without limitation such terms and conditions relating to certain use restrictions, limited license, warranty and limitation of liability, and nothing in this document shall be deemed to be Documentation, as that term is set forth in such terms and conditions of sale. Nothing in this document shall be construed as any representation by 10x Genomics that it currently or will at any time in the future offer or in any way support any application set forth herein.

Contact:

support@10xgenomics.com 10x Genomics 7068 Koll Center Parkway Suite 401 Pleasanton, CA 94566 USA

