

CG00020 Rev D

USER GUIDE

Chromium Controller Specifications

FOR USE WITH

Chromium Controller & Accessory Kit, PN-1000202

Chromium Controller & Accessory Kit, PN-1000204

Notices

Document Number

CG00020 • Rev D

Legal Notices

© 2020 10x Genomics, Inc. (10x Genomics). All rights reserved. Duplication and/or reproduction of all or any portion of this document without the express written consent of 10x Genomics, is strictly forbidden. Nothing contained herein shall constitute any warranty, express or implied, as to the performance of any products described herein. Any and all warranties applicable to any products are set forth in the applicable terms and conditions of sale accompanying the purchase of such product. 10x Genomics provides no warranty and hereby disclaims any and all warranties as to the use of any third-party products or protocols described herein. The use of products described herein is subject to certain restrictions as set forth in the applicable terms and conditions of sale accompanying the purchase of such product. A non-exhaustive list of 10x Genomics' marks, many of which are registered in the United States and other countries can be viewed at: www.10xgenomics.com/trademarks. 10x Genomics may refer to the products or services offered by other companies by their brand name or company name solely for clarity, and does not claim any rights in those third party marks or names. 10x Genomics products may be covered by one or more of the patents as indicated at: www.10xgenomics.com/patents. The use of products described herein is subject to 10x Genomics Terms and Conditions of Sale, available at www.10xgenomics.com/legal-notices, or such other terms that have been agreed to in writing between 10x Genomics and user. All products and services described herein are intended FOR RESEARCH USE ONLY and NOT FOR USE IN DIAGNOSTIC PROCEDURES.

Instrument & Licensed Software Updates Warranties

Updates to existing Instruments and Licensed Software may be required to enable customers to use new or existing products. In the event of an Instrument failure resulting from an update, such failed Instrument will be replaced or repaired in accordance with the 10x Limited Warranty, Assurance Plan or service agreement, only if such Instrument is covered by any of the foregoing at the time of such failure. Instruments not covered under a current 10x Limited Warranty, Assurance Plan or service agreement will not be replaced or repaired.

Support

Email: support@10xgenomics.com

10x Genomics

7068 Koll Center Parkway

Suite 401

Pleasanton, CA 94566 USA

Document
Revision
Summary

Document Number	CG00020
Title	Chromium Controller Specifications
Revision	Rev C to Rev D
Revision Date	December 2020

Specific Changes:

- Updated to include Chromium Controller & Accessory Kit part numbers on page 4.

Chromium Controller & Accessories

Chromium Controller & Next GEM Accessory Kit, PN-1000202/PN-1000204

- Chromium Controller, PN 120270
- Chromium Next GEM Accessory Kit, PN-1000142 (store at ambient temperature)

Next GEM Accessory Kit Items	#	PN (Item)
Chromium Next GEM Test Chip	1	2000199
Chromium Next GEM Secondary Holder	2	3000332
Vortex Adaptor	3	330002
Chip Gaskets, 2 Pack	1	3000072
Magnetic Separator	1	230003
Strip Vortexer	2	230002
Cable USB A MALE - B MALE 1M BLK	1	3000173
Power Cord	1	340001(North America) <i>(Item PN is region specific)</i>

1. Safety and Compliance Information

1.1. Regulatory

The Chromium Controller has been designed, tested, and certified to be in compliance with the following standards:



- UL 61010-1:2012 and CAN/CSA C22.2 No. 61010-1-12 with a cTUVus mark to indicate that the product has been tested and certified to Canadian and US standards by TUV Rheinland and can be legally installed in those countries.
- IEC/EN 61010-1:2010 (3rd Edition): Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory use.
- EN 61326-1:2013: Electrical Equipment for Measurement, Control and Laboratory Use. EMC Requirements.



- The RCM mark indicates an electrical product complies with all the requirements of the electrical and EMC regulations of Australia and New Zealand in accordance with AS/NZS Standards.



- CE Mark indicates that the Chromium Controller assembly is covered by a Declaration of Conformity, and has been declared in conformity with the provisions of all applicable directives in the European Union.



- RoHS Directive (2011/65/EU): Restriction of Hazardous Substances in Electrical and Electronic Equipment
- WEEE Directive (2012/19/EU): Waste Electrical and Electronic Equipment
- FCC Part 15 Class A. NOTE: This equipment has been tested and found to comply with the limits for a Class A digital device, pursuant to part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference when the equipment is operated in a commercial environment. This equipment generates, uses, and can radiate radio frequency energy and, if not installed and used in accordance with the instruction manual, may cause harmful interference to radio communications. Operation of this equipment in a residential area is likely to cause harmful interference in which case the user will be required to correct the interference at his own expense. This device complies with part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) This device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.
- ICES-003 (Canada): This Class A digital apparatus complies with Canadian ICES-003. Cet appareil numérique de la classe A est conforme à la norme NMB-003 du Canada.

SPECIFICATIONS

1.2. Instrument Safety

Before operating this instrument please ensure that everyone involved with the operation of the instrument has:

- Received instruction in general safety practices for laboratories
- Received instruction in specific safety practices for the instrument
- Read and understood all related Safety Data Sheet (SDS) documents

Precautions are illustrated in the following way:



WARNING: Indicates a potentially hazardous situation, which, if not avoided, may result in minor or moderate injury. It may also be used to alert against unsafe practices.



CAUTION: Indicates a potentially hazardous situation, which, if not avoided, could result in death or serious injury.

2. Product Specifications

Weight	12.5 lb (5.6 kg)
Size (W X D X H)	7.9 x 10.3 x 6.4 in (20 x 26.3 x 16.4 cm)
Electrical Requirements	Nominal 100-240 Vac, 50-60 Hz, 25 W, from a standard 3-prong wall receptacle that includes a safety ground pin Maximum 90-264V operational range (+/- 10% of Nominal) Overvoltage Category II (Standard Receptacle)
Pollution Degree	2 (Indoor Use Only)
Ventilation Requirement	Minimum 4 in (10 cm) Around All Sides
Usable Temperature Range	18-28°C (64-82°F)
Humidity	85% Max (Non-Condensing)
Altitude	0 - 6560 ft (0 - 2000 m)

3. Initial Installation Notes

To operate the Chromium Controller properly and obtain maximum performance, locate the unit with the following conditions:

- Remove all transportation packaging materials and tape.
- Choose a location not subject to direct sunlight, as direct sunlight could lead to inadequate cooling conditions or rapid evaporation of reagents.
- Always keep a minimum of **4 in (10 cm)** of clearance around unit to allow free air circulation.
- Leave a minimum of **6 in (15 cm)** directly in front of the instrument for unobstructed movement of the tray door.



WARNING: The tray door is capable of moving an object that is in its opening path. If near the edge of the workspace, the object could fall and create a hazard.

- Choose a location on a level sturdy laboratory bench that is not subject to movement or vibration.
- Connect the Chromium Controller to a properly grounded outlet of proper voltage. Do not position the instrument in such a way that it would be difficult to disconnect the power source. After installation, it should be possible to 1) reach the plug at the wall receptacle to safely disconnect power, or 2) reach the detachable power cord at the rear of the instrument. These areas should remain accessible for the potential need to disconnect the instrument, e.g., for servicing.



CAUTION: Ensure that the ground is reliably connected before plugging the Chromium Controller's power cord into the power source (receptacle). Grounding is required to prevent electric shock. If the power source is not grounded, qualified personnel must first install a reliable safety ground.



CAUTION: Avoid using the Chromium Controller in a manner not specified by 10x Genomics®. The Chromium Controller has been designed to protect the user. If used improperly, the intended user protections can be impaired.

- Before running Chips in the Chromium Controller, the Readiness Test should be run according to instructions provided in *Chromium Controller Readiness Test – CG00025*.

4. Maintenance

4.1. Interior

The Chromium™ Controller tray and door areas have been designed to catch and contain drips and small volume of liquid spills. Occasionally, use a soft lab towel to clean these areas with a mixture of mild detergent and distilled water. For deeper, more thorough cleaning, it is acceptable to use a **5-10% Bleach** solution followed by a **70% ethanol** wash.

NOTE: Do not use acetone or other harsh solvents as these may remove the colored markings in the tray. Apply all standard safety practices when using cleaners, and dispose of any generated waste in a responsible manner.

4.2. Exterior

The exterior of the Chromium Controller should always be kept clean and free of dust and debris that may affect its function and/or cooling efficiency. Generally, the exterior finish can be wiped down using a mixture of mild detergent and distilled water applied to a slightly dampened lab towel. As an added precaution it is recommended that the instrument be unplugged from the power source before beginning any cleaning process.

4.3. Service



WARNING: Electrical shock hazard. Do not open the Chromium Controller. There are no user serviceable parts inside. Refer all servicing to qualified 10x Genomics® service personnel.

Servicing is required when the Chromium Controller has been damaged in any way, e.g., a power entry module or plug is damaged, liquid was spilled into, or objects fell into the instrument, the instrument does not operate properly, or has been dropped.

Use only the power cord supplied with the Chromium Controller. Do not replace it with a non-approved power cord as it may be inadequately rated to handle the electrical loads.

If replacing the externally accessible fuses in the power entry module becomes necessary, use only **certified** (EN60127 Sheet 5) **5 x 20 mm sized fuses rated T1AH, 250V Slow-Blow** or equivalent.

When returning a Chromium Controller for repair, please take steps to ensure that the instrument has been decontaminated so as not to pose a hazard for 10x Genomics® service personnel.

4.4. Moving the Instrument

If the Chromium Controller will be transported in a vehicle or shipped, it should be repackaged according to the *Chromium Controller Shipping Guide* – CG000058. The Readiness Test should be re-run (see *Chromium Controller Readiness Test* – CG000025 for instructions) after relocating the instrument.

4.5. Environmental Requirements

It is the design intent of the Chromium™ Controller that it be used in a typical indoor laboratory environment. The instrument's operating temperature is **18-28°C (64-82°F)**; too low or too high temperature conditions will affect the sensitive reagents used with the instrument.