CG000399 Rev A

USER GUIDE

Chromium Next GEM Single Cell 3' LT Reagent Kits v3.1 (Dual Index)



FOR USE WITH

Chromium Next GEM Single Cell 3' LT Kit v3.1, 4 rxns PN-1000325 Chromium Next GEM Chip L Single Cell Kit, 16 rxns PN-1000321 Dual Index Kit TT Set A, 96 rxns PN-1000215



Next GEM reagents are specific to Next GEM products and should not be used interchangeably with non-Next GEM reagents.

10xGenomics.com

Notices

Document Number

CG000399 • Rev A

Legal Notices

© 2021 10x Genomics, Inc. (10x Genomics). All rights reserved. Duplication and/or reproduction of all or any portion of this document without the express written consent of 10x Genomics, is strictly forbidden. Nothing contained herein shall constitute any warranty, express or implied, as to the performance of any products described herein. Any and all warranties applicable to any products are set forth in the applicable terms and conditions of sale accompanying the purchase of such product. 10x Genomics provides no warranty and hereby disclaims any and all warranties as to the use of any third-party products or protocols described herein. The use of products described herein is subject to certain restrictions as set forth in the applicable terms and conditions of sale accompanying the purchase of such product. A non-exhaustive list of 10x Genomics' marks, many of which are registered in the United States and other countries can be viewed at: www.10xgenomics.com/trademarks. 10x Genomics may refer to the products or services offered by other companies by their brand name or company name solely for clarity, and does not claim any rights in those third party marks or names. 10x Genomics products may be covered by one or more of the patents as indicated at: www.10xgenomics.com/patents. The use of products described herein is subject to 10x Genomics Terms and Conditions of Sale, available at www.10xgenomics.com/legal-notices, or such other terms that have been agreed to in writing between 10x Genomics and user. All products and services described herein are intended FOR RESEARCH USE ONLY and NOT FOR USE IN DIAGNOSTIC PROCEDURES.

Instrument & Licensed Software Updates Warranties

Updates to existing Instruments and Licensed Software may be required to enable customers to use new or existing products. In the event of an Instrument failure resulting from an update, such failed Instrument will be replaced or repaired in accordance with the 10x Limited Warranty, Assurance Plan or service agreement, only if such Instrument is covered by any of the foregoing at the time of such failure. Instruments not covered under a current 10x Limited Warranty, Assurance Plan or service agreement will not be replaced or repaired.

Support

Email: support@10xgenomics.com 10x Genomics 6230 Stoneridge Mall Road Pleasanton, CA 94588 USA

Document Number	CG000399	
Title	Chromium Next GEM Single Cell 3' LT Reagent Kits v3.1 (Dual Index) User Guide	
Revision	Rev A	
Revision Date	March 2021	
	Title Revision	

Table of Contents

Introduction	5
Chromium Next GEM Single Cell 3' LT Reagents Kits	6
10x Genomics Accessories	9
Recommended Thermal Cyclers	9
Additional Kits, Reagents & Equipment	10
Protocol Steps & Timing	12
Stepwise Objectives	13
Tips & Best Practices	16
Step 1	23
GEM Generation & Barcoding	24
1.1 Prepare Master Mix	25
1.2 Load Chromium Next GEM Chip L	28
1.3 Run the Chromium Controller	29
1.4 Transfer GEMs	30
1.5 GEM-RT Incubation	30
Step 2	31
Post GEM-RT Cleanup & cDNA Amplification	32
2.1 Post GEM-RT Cleanup – Dynabeads	33
2.2 cDNA Amplification	35
2.3 cDNA Cleanup – SPRIselect	36
2.4 Post cDNA Amplification QC & Quantification	37
Step 3	38
3' Gene Expression Library Construction	39
3.1 Fragmentation, End Repair & A-tailing	41
3.2 Post Fragmentation, End Repair & A-tailing Double Sided Size Selection – SPRIselect	42
3.3 Adaptor Ligation	43
3.4 Post Ligation Cleanup – SPRIselect	44
3.5 Sample Index PCR	45
3.6 Post Sample Index PCR Double Sided Size Selection – SPRIselect	46
3.7 Post Library Construction QC	47
Sequencing	48

Troubleshooting	52
GEMs	53
Chromium Controller Errors	55
Appendix	56
Post Library Construction Quantification	
Agilent TapeStation Traces	58
LabChip Traces	59
Oligonucleotide Sequences	60

TOC

Introduction

Chromium Next GEM Single Cell 3' LT Reagent Kits v3.1 (Dual Index) 10x Genomics Accessories Recommended Thermal Cyclers Additional Kits, Reagents & Equipment Protocol Steps & Timing Stepwise Objectives

Chromium Next GEM Single Cell 3' LT Reagents Kits

Chromium Next GEM Single Cell 3' LT Kit v3.1, 4 rxns PN-1000325

n romium ext GEM ngle Cell 3' LT		Library Construction Kit	
EM Kit v3.1	# PN		#
RT Reagent B	1 2000165	Fragmentation Enzyme	1
RT Enzyme C	1 2000102	Fragmentation Buffer	1
Template Switch Oligo	1 3000228	Ligation Buffer	1
Reducing Agent B	1 2000087	🛑 DNA Ligase	1
Cleanup Buffer	1 2000088	🥚 Adaptor Oligos	1
cDNA Primers	1 2000089		
) Amp Mix	1 2000103		
Genomics.com	10x	10xGenomics.com	

(store at -80°C)

	Chromium Next GEM Single Cell 3' LT v3.1 Gel Beads	# PN		
This gel bead kit is specific for the Single Cell 3' LT protocol /assay and should not be used interchangeably with other Single Cell 3' assays.	Single Cell 3' LT v3.1 Gel Beads (4 rxns)	1 2000360		
	10xGenomics.com	10× genomics		
Dynabeads™ MyOne™ SILANE PN-2000048 (store at 4°C)				

PN Dynabeads MyOne 1 2000048 SILANE

Dual Index Kit TT Set A, 96 rxns PN-1000215 (store at -20°C)

Dual Index Kit TT Set A	
	# PN
Dual Index Plate TT Set A	1 3000431

Chromium Next GEM Chip L Single Cell Kit, 16 rxns PN-1000321 (store at ambient temperature)

Chromium Partitioning Oil	# PN	Chromium Recovery Agent	#	PN
Partitioning Oil	2 2000190	Recovery Agent	2	220016
Chromium Next GEM Chip L & Ga		# PN		
Chromi	ium Next GEM Cl	hip L 2 2000414		
Chip Ga	asket, 2-pack	1 3000072		
10xGenomics.com				10x genomics

10x Genomics Accessories

Product	Part Number (Kit)	Part Number (Item)
10x Vortex Adapter	120251	330002
10x Magnetic Separator	120250	230003
Chromium Next GEM Secondary Holder	1000142	3000332

Recommended Thermal Cyclers

Thermal cyclers used must support uniform heating of 100 μl emulsion volumes.

Supplier	Description	Part Number
BioRad	C1000 Touch Thermal Cycler with 96-Deep Well Reaction Module	1851197
Eppendorf	MasterCycler Pro (discontinued)	North America 950030010 International 6321 000.019
Thermo Fisher Scientific	Veriti 96-Well Thermal Cycler	4375786

Additional Kits, Reagents & Equipment

The items in the table below have been validated by 10x Genomics and are highly recommended for the Single Cell 3' protocols. Substituting materials may adversely affect system performance. This list may not include some standard laboratory equipment.

Supplier	Description		Part Number (US)
Plastics			
Eppendorf	PCR Tubes 0.2 ml 8-tube strips DNA LoBind Tubes, 1.5 ml DNA LoBind Tubes, 2.0 ml	, 1.5 ml	
USA Scientific	TempAssure PCR 8-tube strip	Thermo Fisher Scientific PCR	1402-4700
Thermo Fisher Scientific	MicroAmp 8-Tube Strip, 0.2 ml MicroAmp 8 -Cap Strip, clear	8-tube strips.	N8010580 N8010535
Rainin	Tips LTS 200UL Filter RT-L200FLR Tips LTS 1ML Filter RT-L1000FLR Tips LTS 20UL Filter RT-L10FLR		30389240 30389213 30389226
Kits & Reagents			
Thermo Fisher Scientific	Nuclease-free Water Low TE Buffer (10 mM Tris-HCl pH 8.0, 0.1 mM E	DTA)	AM9937 12090-015
Millipore Sigma	Ethanol, Pure (200 Proof, anhydrous)		E7023-500ML
Beckman Coulter	SPRIselect Reagent Kit		B23318
Bio-Rad	10% Tween 20		1662404
Ricca Chemical Company	Glycerin (glycerol), 50% (v/v) Aqueous Solution		3290-32
Qiagen	Qiagen Buffer EB		19086
Equipment			
VWR	Vortex Mixer Divided Polystyrene Reservoirs VWR Mini Centrifuge (alternatively, use any equivalent mini centrifuge)		10153-838 41428-958 76269-066
Eppendorf	Eppendorf ThermoMixer C Eppendorf SmartBlock 1.5 ml, Thermoblock for 24 reaction vessel (alternatively, use a temperature-controlled Heat Block)		5382000023 5360000038
Rainin	Pipet-Lite Multi Pipette L8-50XLS+ Pipet-Lite Multi Pipette L8-200XLS+ Pipet-Lite Multi Pipette L8-10XLS+ Pipet-Lite Multi Pipette L8-20XLS+ Pipet-Lite LTS Pipette L-2XLS+ Pipet-Lite LTS Pipette L-10XLS+ Pipet-Lite LTS Pipette L-20XLS+ Pipet-Lite LTS Pipette L-200XLS+ Pipet-Lite LTS Pipette L-200XLS+ Pipet-Lite LTS Pipette L-1000XLS+		17013804 17013805 17013802 17013803 17014393 17014388 17014392 17014384 17014391 17014382

Additional Kits, Reagents & Equipment

The items in the table below have been validated by 10x Genomics and are highly recommended for the Single Cell 3' protocols. Substituting materials may adversely affect system performance. This list may not include some standard laboratory equipment.

Supplier	Description		Part Number (US)
Quantification & Quality Control			
Agilent	2100 Bioanalyzer Laptop Bundle High Sensitivity DNA Kit 4200 TapeStation High Sensitivity D1000 ScreenTape/Reagents High Sensitivity D5000 ScreenTape/Reagents	Choose Bioanalyzer, TapeStation, LabChip, Fragment Analyzer or Qubit	G2943CA 5067-4626 G2991AA 5067-5584/5067-5585 5067-5592/5067-5593
Thermo Fisher Scientific	Qubit 4.0 Flourometer Qubit dsDNA HS Assay Kit	based on availability & preference.	Q33226 Q32854
Advanced Analytical	Fragment Analyzer Automated CE System - 12 cap Fragment Analyzer Automated CE System - 48/96 cap High Sensitivity NGS Fragment Analysis Kit		FSv2-CE2F FSv2-CE10F DNF-474
PerkinElmer	LabChip GX Touch HT Nucleic Acid Analyzer DNA High Sensitivity Reagent Kit		CLS137031 CLS760672
KAPA Biosystems	KAPA Library Quantification Kit for Illumina Platforms		KK4824

Protocol Steps & Timing

Day	Steps	Timing Stop & Store
	Cell Preparation & Labeling	
1 h	Dependent on Cell Type	~1-2 h
	Step 1 – GEM Generation & Barcoding	
	 Prepare Reaction Mix Load Chromium Next GEM Chip G Run the Chromium Controller Transfer GEMs 	20 min 10 min 18 min 3 min
4h	1.5 GEM-RT Incubation	55 min
	Step 2 – Post GEM-RT Cleanup & cDNA Amplification	
	 2.1 Post GEM RT-Cleanup – Dynabead 2.2 cDNA Amplification 2.3 cDNA Cleanup – SPRIselect 2.3A Pellet Cleanup 	45 min 40 min
	2.3B Transferred Supernatant Cleanup 2.4 cDNA QC & Quantification	20 min
6h	Step 3 – 3' Gene Expression Library Construction	
	 3.1 Fragmentation, End Repair & A-tailing 3.2 Post Fragmentation, End Repair & A-tailing Double Sided Size Selection – SPRIselect 	45 min 30 min
8 h	 3.3 Adaptor Ligation 3.4 Post Ligation Cleanup- SPRIselect 3.5 Sample Index PCR 3.6 Post Sample Index PCR Double Sided Size Selection- SPRIselect 3.7 Post Library Construction QC 	25 min 20 min 40 min 30 min 50 min 50 min

Stepwise Objectives




The Chromium Single Cell Gene Expression Solution upgrades short read sequencers to deliver a scalable microfluidic platform for assessing cell surface protein expression combined with 3' digital gene expression of the same single cell by profiling 100-1,000 individual cells per sample. A pool of 9,216 10x Barcodes are sampled separately to index each cell's transcriptome. It is done by partitioning thousands of cells into nanoliter-scale Gel Beads-in-emulsion (GEMs), where all generated DNA molecules share a common 10x Barcode. Libraries are generated and sequenced from the DNA molecules and 10x Barcodes are used to associate individual reads back to the individual partitions.

This document outlines the low throughput (LT) protocol for generating Single Cell 3' Gene Expression.

Single Cell 3' LT v3.1 Gel Beads




In addition to the poly(dT) primer that enables the production of barcoded, full-length cDNA from poly-adenylated mRNA, the Single Cell 3' LT v3.1 Gel Beads also include two additional primer sequences (Capture Sequence 1 and Capture Sequence 2), that enable capture and priming of Feature Barcode technology compatible targets or analytes of interest. Only the poly(dT) primers are used in this protocol for generating Single Cell 3' Gene Expression libraries.



This gel bead is specific for the Single Cell 3' LT protocol /assay and should not be used interchangeably with other Single Cell 3' assays.

Step 1 GEM Generation & Barcoding

GEMs are generated by combining barcoded Single Cell 3' LT v3.1 Gel Beads, a Master Mix with cell surface protein labeled cells, and Partitioning Oil onto Chromium Chip L. To achieve single cell resolution, cells are delivered at a limiting dilution, such that the majority (~90-99%) of generated GEMs contain no cell, while the remainder largely contain a single cell.




Step 1 GEM Generation & Barcoding

Immediately following GEM generation, the Gel Bead is dissolved, primers are released, and any co-partitioned cell is lysed. Primers containing:


- an Illumina TruSeq Read 1 (read 1 sequencing primer)
- 16 nt 10x Barcode
- 12 nt unique molecular identifier (UMI)
- 30 nt poly(dT) sequence

are mixed with the cell lysate and a Master Mix containing reverse transcription (RT) reagents. Incubation of the GEMs produces barcoded, full-length cDNA from polyadenylated mRNA.



Step 2 Post GEM-RT Cleanup & cDNA Amplification


After incubation, GEMs are broken and pooled fractions are recovered. Silane magnetic beads are used to purify the first-strand cDNA from the post GEM-RT reaction mixture, which includes leftover biochemical reagents and primers. Barcoded, full-length cDNA is amplified via PCR to generate sufficient mass for library construction.



Step 3 3' Gene Expression Library Construction




Enzymatic fragmentation and size selection are used to optimize the cDNA amplicon size. P5, P7, i7 and i5 sample indexes, and TruSeq Read 2 (read 2 primer sequence) are added via End Repair, A-tailing, Adaptor Ligation, and PCR. The final libraries contain the P5 and P7 primers used in Illumina amplification.



Step 4 Sequencing

A Chromium Single Cell 3' Gene Expression Dual Index library comprises standard Illumina paired-end constructs which begin and end with P5 and P7. The 16 bp 10x Barcode and 12 bp UMI are encoded in Read 1, while Read 2 is used to sequence the cDNA fragment. i7 and i5 index sequences are incorporated as the sample index reads. TruSeq Read 1 and TruSeq Read 2 are standard Illumina sequencing primer sites used in paired-end sequencing.

Illumina sequencer compatibility, sample indices, library loading and pooling for sequencing are summarized in step 4.



Chromium Single Cell 3' Gene Expression Dual Index Library

See Appendix for Oligonucleotide Sequences

Tips & Best Practices



Emulsion-safe Plastics

• Use validated emulsion-safe plastic consumables when handling GEMs as some plastics can destabilize GEMs. Refer to Additional Kits, Reagents & Equipments table for validated plastics.

Cell Concentration



- Recommended starting point is to load ~1,667 cells per reaction, resulting in recovery of ~500 cells, and a multiplet rate of ~4%. The optimal input cell concentration is 100-600 cells/µl.
- Refer to Cell Dilution Guidelines for preparing cells.
- The presence of dead cells in the suspension may also reduce the recovery rate. Consult the 10x Genomics Single Cell Protocols Cell Preparation Guide and the Guidelines for Optimal Sample Preparation flowchart (Documents CG00053 and CG000126 respectively) for more information on preparing cells.

Multiplet Rate (%)	# of Cells Loaded	# of Cells Recovered
~0.8%	~333	~100
~1.5%	~667	~200
~2.3%	~1,000	~300
~3.0%	~1,333	~400
~3.8%	~1,667	~500
~4.6%	~2,000	~600
~5.3%	~2,333	~700
~6.1%	~2,667	~800
~6.8%	~3,000	~900
~8.0%	~3,333	~1,000

General Reagent	 Fully thaw and thoroughly mix reagents before use.
Handling	 Keep all enzymes and Master Mixes on ice during setup and use. Promptly move reagents back to the recommended storage.
	 Calculate reagent volumes with 10% excess of 1 reaction values.
	Cover Partitioning Oil tubes and reservoirs to minimize evaporation.
	• If using multiple chips, use separate reagent reservoirs for each chip during loading.
	 Thoroughly mix samples with the beads during bead-based cleanup steps.
50% Glycerol Solution	 Purchase 50% glycerol solution from Ricca Chemical Company, Glycerin (glycerol), 50% (v/v) Aqueous Solution, PN-3290-32.
	Prepare 50% glycerol solution:
	i. Mix an equal volume of water and 99% Glycerol, Molecular Biology Grade.
	ii. Filter through a 0.2 μ m filter.
	iii. Store at –20°C in 1-ml LoBind tubes. 50% glycerol solution should be equilibrated to room temperature before use.
Pipette Calibration	Follow manufacturer's calibration and maintenance schedules.
	 Pipette accuracy is particularly important when using SPRIselect reagents.
Chromium Next GEM Chip Handling	 Minimize exposure of reagents, chips, and gaskets to sources of particles and fibers, laboratory wipes, frequently opened flip-cap tubes, clothing that sheds fibers, and dusty surfaces.
	• After removing the chip from the sealed bag, use in \leq 24 h.
	• Execute steps without pause or delay, unless indicated. When multiple chips are to be used, load, run, and collect the content from one chip before loading the next.
	 Fill all unused input wells in rows labeled 1, 2, and 3 on a chip with an appropriate volume of 50% glycerol solution before loading the used wells. DO NOT add glycerol to the wells in the bottom NO FILL row.
	 Avoid contacting the bottom surface of the chip with gloved hands and other surfaces. Frictional charging can lead to inadequate priming of the channels, potentially leading to either clogs or wetting failures.
	• Minimize the distance that a loaded chip is moved to reach the Chromium Controller.
	 Keep the chip horizontal to prevent wetting the gasket with oil, which depletes the input volume and may adversely affect the quality of the resulting emulsion.

Tips & Best Practices

Chromium Next GEM Secondary	Chromium Next GEM Secondary Holders encase Chromium Next GEM Chips. Sliders Sliders
Holders	• The holder lid flips over to become a stand, holding the chip at 45 degrees for optimal recovery well content removal.
	• Squeeze the black sliders on the back side of the holder together to unlock the lid and return the holder to a flat position.
Chromium Next	Align notch on the chip (upper left corner) and the holder. Chromium Next GEM Secondary Holder
GEM Chip & Holder Assembly	• Insert the left-hand side of the chip under the guide. Depress the right-hand side of the chip until the spring-loaded clip engages.
	Close the lid before dispensing reagents into the wells.
Chromium Next GEM	Place the assembled chip and holder flat on the bench with the lid closed.
Chip Loading	 Dispense at the bottom of the wells without introducing bubbles.
	• When dispensing Gel Beads into the chip, wait for the remainder to drain into the bottom of the pipette tips and dispense again to ensure complete transfer.
	Refer to Load Chromium Next GEM Chip L for specific instructions.
Gel Bead Handling	Use one tube of Gel Beads per sample. DO NOT puncture the foil seals of tubes not used at the time.
	Equilibrate the Gel Beads strip to room temperature before use.
	 Store unused Gel Beads at -80°C and avoid more than 12 freeze-thaw cycles. D0 N0T store Gel Beads at -20°C.
	• Snap the tube strip holder with the Gel Bead strip into a 10x Vortex Adapter. Vortex 30 sec .
	 Centrifuge the Gel Bead strip for ~5 sec after removing from the holder. Confirm there are no bubbles at the bottom of the tubes and the liquid levels look even. Place the Gel Bead strip back in the holder and secure the holder lid.
	• If the required volume of beads cannot be recovered, place the pipette tips against the sidewalls and slowly dispense the Gel Beads back into the tubes. DO NOT introduce bubbles into the tubes and verify that the pipette tips contain no leftover Gel Beads. Withdraw the full volume of beads again by pipetting slowly.

10x Gasket Attachment

- After reagents are loaded, attach the gasket by holding the tongue (curved end, to the right) and hook it on the left-hand tabs of the holder. Gently pull the gasket toward the right and hook it on the two right-hand tabs.
 - DO NOT touch the smooth side of the gasket. DO NOT press down on the top of the gasket after attachment.
 - Keep the assembly horizontal to avoid wetting the gasket with Partitioning Oil.



10x Magnetic Separator


 Offers two positions of the magnets (high and low) relative to a tube, depending on its orientation. Flip the magnetic separator over to switch between high (magnet•High) or low (magnet•Low) positions.



• If using MicroAmp 8-Tube Strips, use the high position (magnet•High) only throughout the protocol.


Magnetic Bead Cleanup Steps

- During magnetic bead based cleanup steps that specify waiting "until the solution clears", visually confirm clearing of solution before proceeding to the next step. See adjacent panel for an example.
- The time needed for the solution to clear may vary based on specific step, reagents, volume of reagents etc.



SPRIselect Cleanup & Size Selection

- After aspirating the desired volume of SPRIselect reagent, examine the pipette tips before dispensing to ensure the correct volume is transferred.
- Pipette mix thoroughly as insufficient mixing of sample and SPRIselect reagent will lead to inconsistent results.
- Use fresh preparations of 80% Ethanol.



Enzymatic Fragmentation

 Ensure enzymatic fragmentation reactions are prepared on ice and then loaded into a thermal cycler pre-cooled to 4°C prior to initiating the Fragmentation, End Repair, and A-tailing incubation steps.

Sample Indices in Sample Index PCR

- Choose the appropriate sample index sets to ensure that no sample indices overlap in a multiplexed sequencing run.
- Verify and use the specified index plate only. DO NOT use the plates interchangebaly.
- Each well in the Dual Index Plate contains a unique i7 and a unique i5 oligonucleotide.

Index Hopping Mitigation

Index hopping can impact pooled samples sequenced on Illumina sequencing platforms that utilize patterned flow cells and exclusion amplification chemistry. To minimize index hopping, follow the guidelines listed below.

- Remove adapters during cleanup steps.
- Ensure no leftover primers and/or adapters are present when performing post-Library Construction QC.
- Store each library individually at 4°C for up to 72 h or at -20°C for long-term storage.
 DO NOT pool libraries during storage.
- Pool libraries prior to sequencing. An additional 1.0X SPRI may be performed for the pooled libraries to remove any free adapters before sequencing.
- Hopped indices can be computationally removed from the data generated from single cell dual index libraries.

Step 1

GEM Generation & Barcoding

- **1.1** Prepare Single Cell Master Mix
- 1.2 Load Chromium Next GEM Chip L
- **1.3** Run the Chromium Controller
- 1.4 Transfer GEMs
- 1.5 GEM-RT Incubation

Step 1

1.0 GEM Generation & Barcoding



GET STARTED!					
Action		ltem	10x PN	Preparation & Handling	Storage
Equilibrate to Room Temperature		Single Cell 3' LT v3.1 Gel Beads	2000324	Equilibrate to room temperature 30 min before loading the chip.	-80°C
	•	RT Reagent B	2000165	Vortex, verify no precipitate, centrifuge briefly.	–20°C
		Template Switch Oligo	3000228	Centrifuge briefly, resuspend in 80 μ L Low TE Buffer. Vortex 15 sec at maximum speed, centrifuge briefly, leave at room temperature for \geq 30 min. After resuspension, store at -80°C.	-20°C
	\bigcirc	Reducing Agent B	2000087	Vortex, verify no precipitate, centrifuge briefly.	–20°C
Place on Ice	•	RT Enzyme C	2000102	Centrifuge briefly before adding to the mix.	–20°C
Obtain		Partitioning Oil	2000190	-	Ambient
		Chromium Next GEM Chip L	2000414	-	Ambient
		10x Gasket	3000072	See Tips & Best Practices.	Ambient
		Chromium Next GEM Secondary Holder	3000332	See Tips & Best Practices.	Ambient
		10x Vortex Adapter	330002	See Tips & Best Practices.	Ambient
		50% glycerol solution	-	See Tips & Best Practices.	-



required in the Chromium Controller or the Chromium Single Cell Controller used for this protocol.

1.1 Prepare Master Mix

a. Prepare Master Mix on ice. Pipette mix 15x and centrifuge briefly.

Master Mix Add reagents in the order listed	PN	1Χ (μl)	4X + 10% (μl)
RT Reagent B	2000165	18.8	82.7
Template Switch Oligo	3000228	2.4	10.6
O Reducing Agent B	2000087	2.0	8.8
RT Enzyme C	2000102	8.7	38.3
Total	-	31.9	140.4

b. Add **31.9 µl** Master Mix into each tube of a PCR 8-tube strip on ice.



Assemble Chromium Next GEM Chip L

After removing the chip from the sealed bag, use the chip in \leq 24 h.

TIPS

See Tips & Best Practices for chip handling instructions.

- Align notch on the chip (upper left corner) and the holder.
- Insert the left-hand side of the chip under the guide. Depress the righthand side of the chip until the springloaded clip engages.
- Close the lid before dispensing reagents into the wells.
- The assembled chip is ready for loading the indicated reagents. Refer to step 1.2 for reagent volumes and loading order.







For GEM generation, load the indicated reagents only in the specified rows, starting from row labeled 1, followed by rows labeled 2 and 3. DO NOT load reagents in the bottom row labeled NO FILL. See step 1.2 for details.




Cell Dilution Guidelines



Step 1

The optimal cell loading concentration for this protocol is ~100-600 cells/µl. Dilute cells using only compatible cell media/buffer (minimize carryover of media that can destabilize GEMs & impact GEM-RT reaction). Always mix cells gently and thoroughly before taking an aliquot. See dilution overview and examples below.



Nuclease-free water vol. per rxn (µl)

41.1 µl

39.3 µl



Cell Suspension Volume Calculator Table


(for step 1.2 of Chromium Next GEM Single Cell 3' Low Throughput v3.1 protocol)

Volume of Cell Suspension Stock per reaction (µl) | Volume of Nuclease-free Water per reaction (µl)



The optimal cell loading concentration for this protocol is **~100-600 cells/ul**. Ensure cell counting is conducted in the optimal cell counting concentration and cell stock is diluted to the optimal cell loading concentration prior to adding the single cell suspension volume to the Master Mix for chip loading. DO NOT add nuclease-free water directly to single cell suspension. Add nuclease-free water to the Master Mix. Refer to step 1.2b.

Cell Stock	Stock Targeted Cell Recovery									
Concentration (Cells/µl)	100	200	300	400	500	600	700	800	900	1000
100	3.3	6.7	10.0	13.3	16.7	20.0	23.3	26.7	30.0	33.3
100	40.0	36.6	33.3	30.0	26.6	23.3	20.0	16.6	13.3	10.0
200	1.7	3.3	5.0	6.7	8.3	10.0	11.7	13.3	15.0	16.7
200	41.6	40.0	38.3	36.6	35.0	33.3	31.6	30.0	28.3	26.6
300	1.1	2.2	3.3	4.4	5.6	6.7	7.8	8.9	10.0	11.1
300	42.2	41.1	40.0	38.9	37.7	36.6	35.5	34.4	33.3	32.2
400	0.8	1.7	2.5	3.3	4.2	5.0	5.8	6.7	7.5	8.3
400	42.5	41.6	40.8	40.0	39.1	38.3	37.5	36.6	35.8	35.0
500	0.7	1.3	2.0	2.7	3.3	4.0	4.7	5.3	6.0	6.7
500	42.6	42.0	41.3	40.6	40.0	39.3	38.6	38.0	37.3	36.6
(00	0.6	1.1	1.7	2.2	2.8	3.3	3.9	4.4	5.0	5.6
600	42.7	42.2	41.6	41.1	40.5	40.0	39.4	38.9	38.3	37.7
700	0.5	1.0	1.4	1.9	2.4	2.9	3.3	3.8	4.3	4.8
700	42.8	42.3	41.9	41.4	40.9	40.4	40.0	39.5	39.0	38.5
000		0.8	1.3	1.7	2.1	2.5	2.9	3.3	3.8	4.2
800	N/A	42.5	42.1	41.6	41.2	40.8	40.4	40.0	39.6	39.1
000		0.7	1.1	1.5	1.9	2.2	2.6	3.0	3.3	3.7
900	N/A	42.6	42.2	41.8	41.4	41.1	40.7	40.3	40.0	39.6
1000		0.7	1.0	1.3	1.7	2.0	2.3	2.7	3.0	3.3
1000	N/A	42.6	42.3	42.0	41.6	41.3	41.0	40.6	40.3	40.0
Grey b	oxes: Exceed	ds allowable v	volume	Ye	ellow boxes: L	.ow transfer v	olume	Blue	boxes: Optim	al range



When dispensing, raise the pipette tips at the same rate as the liquid is rising, keeping the tips slightly submerged.



	submerged.	
	lycerol solution to each unused well g <8 samples/chip)	Glycerol
 50 μl in e 70 μl in e 	each unused well in row labeled 3 each unused well in row labeled 2 each unused well in row labeled 1 add 50% glycerol solution to the bottom row of NO FILL wells.	45 μl 3, 100 μl 100 μl 100 μl
	use any substitute for 50% glycerol solution.	
b. Prepare Ma	aster Mix + Cell suspension	/
 Refer to t 	he Cell Suspension Volume Calculator Table.	
Add corre	appropriate volume of nuclease-free water to Master Mix. esponding volume of <mark>single cell suspension</mark> to Master Mix. 5 µl <i>in each tube.</i>	Cells
Gently pip	pette mix the cell suspension before adding to the Master Mix.	Master Mix + Water
c. Load Row L	_abeled 1	Master Mix + Sample
 Gently pip 	pette mix the Master Mix + Cell Suspension	10x
	e same pipette tip, dispense 70 μ l Master Mix + Cell Suspensio enter of each well in row labeled 1 without introducing bubble	on into the
d. Prepare Ge	el Beads	Prep Gel Beads
Vortex 30		Lid Lid
 Centrifug 	je the Gel Bead strip for ~5 sec.	
 Confirm t even. 	there are no bubbles at the bottom of the tubes and the liquid	levels are Gel Bead Strip
 Place the 	e Gel Bead strip back in the holder. Secure the holder lid.	
e. Load Row L	Labeled 2	Gel Beads
Puncture	the foil seal of the Gel Bead tubes.	
 Slowly as 	spirate 50 μl Gel Beads.	
•	into the wells in row labeled 2 without introducing bubbles.	50 µl 2 0 0 0 0 0 0
• Wait 30 s	-	50 µl 2 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0

f. Load Row Labeled 3

- Dispense $45~\mu l$ Partitioning Oil into the wells in row~labeled~3 from a reagent reservoir.

Failure to add Partitioning Oil to the top row labeled 3 will prevent GEM generation and can damage the Chromium Controller.

Partitioning Oil



1.2

OC

Load Chromium

NextGEM Chip L

g. Attach GEM Gasket

- Align the notch with the top left-hand corner.
- Ensure the gasket holes are aligned with the wells.
- Avoid touching the smooth surface.

Attach the gasket and run the chip in the Chromium Controller immediately after loading the Partitioning Oil.



1.3 Run the Chromium Controller

- **a.** Press the eject button on the Controller to eject the tray.
- **b.** Place the assembled chip with the gasket in the tray, ensuring that the chip stays horizontal. Press the button to retract the tray.
- **c.** Confirm the Chromium Chip L program on screen. Press the play button.
- d. At completion of the run (~18 min), the Controller will chime. Immediately proceed to the next step.



Firmware Version 4.0 or higher is required in the Chromium Controller or the Chromium Single Cell Controller used for this protocol.






1.4 Transfer GEMs

- **a.** Place a tube strip on ice.
- **b.** Press the eject button of the Controller and remove the chip.
- c. Discard the gasket. Open the chip holder.
 Fold the lid back until it clicks to expose the wells at 45 degrees.
- **d.** Check the volume in rows labeled 1-2. Abnormally high volume in any well indicates a clog.
- e. Slowly aspirate 100 µl GEMs from the lowest points of the recovery wells in the top row labeled 3 without creating a seal between the tips and the bottom of the wells.
- f. Withdraw pipette tips from the wells. GEMs should appear opaque and uniform across all channels. Excess Partitioning Oil (clear) in the pipette tips indicates a potential clog.
- **g.** Over the course of **~20 sec**, dispense GEMs into the tube strip on ice with the pipette tips against the sidewalls of the tubes.
- h. If multiple chips are run back-to-back, cap/ cover the GEM-containing tube strip and place on ice for no more than 1 h.

Expose Wells at 45 Degrees







1.5 GEM-RT Incubation

Use a thermal cycler that can accommodate at least 100 μl volume. A volume of 125 μl is the preferred setting on Bio-Rad C1000 Touch. In alternate thermal cyclers, use highest reaction volume setting.

a. Incubate in a thermal cycler with the following protocol.

Lid Temperature	Reaction Volume	Run Time
53°C	125 µl	~55 min
Step	Temperature	Time
1	53°C	00:45:00
2	85°C	00:05:00
3	4°C	Hold

STOP

b. Store at 4°C for up to 72 h or at –20°C for up to a week, or proceed to the next step.

Step 2

Post GEM–RT Cleanup & cDNA Amplification

- 2.1 Post GEM–RT Cleanup Dynabeads
- 2.2 cDNA Amplification
- 2.3 cDNA Cleanup SPRIselect
- 2.4 cDNA QC & Quantification

2.0 Post GEM-RT Cleanup & cDNA Amplification



A				Chammer
Action	Item	10x PN	Preparation & Handling	Storage
Equilibrate to Room	Reducing Agent B	2000087	Thaw, vortex, verify no precipitate, centrifuge.	–20°C
Temperature	Feature cDNA Primers 2 Verify name & PN	2000097	Vortex, centrifuge briefly.	–20°C
	Dynabeads MyOne SILANE	2000048	Vortex thoroughly (≥30 sec) immediately before adding to the mix.	4°C
	Beckman Coulter SPRIselect Reagent	-	Manufacturer's recommendations.	-
	Agilent Bioanalyzer High Sensitivity Kit If used for QC and quantification	-	Manufacturer's recommendations.	-
	Agilent TapeStation ScreenTape and Reagents If used for QC and quantification	-	Manufacturer's recommendations.	-
	Qubit dsDNA HS Assay Kit If used for QC and quantification	-	Manufacturer's recommendations.	-
Place on ice	O Amp Mix	2000103	Vortex, centrifuge briefly.	–20°C
Thaw at 65°C	Cleanup Buffer	2000088	Thaw for 10 min at 65°C at max speed on a thermomixer. Verify no visible crystals. Cool to room temperature.	–20°C
Obtain	O Recovery Agent	220016	-	Ambient
	Qiagen Buffer EB	-	Manufacturer's recommendations.	-
	Bio-Rad 10% Tween 20	_	Manufacturer's recommendations.	-
	10x Magnetic Separator	230003	-	Ambient
	Prepare 80% Ethanol Prepare 8 ml for 4 reactions	-	Prepare fresh.	Ambient

2.1 Post GEM-RT Cleanup -**Dynabeads**

Step 2

a. Add 125 µl Recovery Agent to each sample at room temperature. DO NOT pipette mix or vortex the biphasic mixture. Wait 2 min.

The resulting biphasic mixture contains Recovery Agent/Partitioning Oil (pink) and aqueous phase (clear), with no persisting emulsion (opaque).

If biphasic separation is incomplete:

Firmly secure the cap on the tube strip, ensuring that no liquid is trapped between the cap and the tube rim. Mix by inverting the capped tube strip 5x, centrifuge briefly, and proceed to step b. DO NOT invert without firmly securing the caps.

A smaller aqueous phase volume indicates a clog during GEM generation.

b. Slowly remove and discard 125 µl Recovery Agent/Partitioning Oil (pink) from the bottom of the tube. DO NOT aspirate any aqueous sample.

c. Prepare Dynabeads Cleanup Mix.

Remove Recovery Agent





- d. Vortex and add 200 µl to each sample. Pipette mix 10x (pipette set to 200 µl).
- e. Incubate 10 min at room temperature (keep caps open). Pipette mix again at ~5 min after start of incubation to resuspend settled beads.









Step 2

f. Prepare Elution Solution I. Vortex and centrifuge briefly.

Elution Solution I Add reagents in the order listed	PN	1X (μl)	10Χ (μl)
Buffer EB	-	98	980
10% Tween 20	-	1	10
O Reducing Agent B	2000087	1	10
Total	-	100	1000



g. At the end of **10 min** incubation, place on a 10x Magnetic Separator•High position (magnet•High) until the solution clears.

A white interface between the aqueous phase and Recovery Agent is normal.

- h. Remove the supernatant (aqueous phase and Recovery Agent).
- i. Add 300 µl 80% ethanol to the pellet while on the magnet. Wait 30 sec.
- j. Remove the ethanol.
- k. Add 200 µl 80% ethanol to pellet. Wait 30 sec.
- l. Remove the ethanol.
- m.Centrifuge briefly. Place on the magnet•Low.
- n. Remove remaining ethanol. Air dry for 1 min.
- o. Remove from the magnet. Immediately add 35.5 µl Elution Solution I.
- p. Pipette mix (pipette set to 30 µl) without introducing bubbles.
- q. Incubate 2 min at room temperature.
- r. Place on the magnet•Low until the solution clears.
- s. Transfer 35 µl sample to a new tube strip.

2.2 cDNA Amplification



a. Prepare cDNA Amplification Mix on ice. Vortex and centrifuge briefly.

	cDNA Amplification Reaction Mix Add reagents in the order listed	PN	1Χ (μl)	4X + 10% (µl)	8X + 10% (μl)
0	Amp Mix	2000047/ 2000103	50	220	440
	cDNA Primers	2000089	15	66	132
	Total	-	65	286	572

b. Add **65 µl** cDNA Amplification Reaction Mix to **35 µl** sample.

c. Pipette mix 15x (pipette set to 90 µl). Centrifuge briefly.

d. Incubate in a thermal cycler with the following protocol.

Lid Temperature	Reaction Volume	Run Time
105°C	100 µl	~30-45 min
Step	Temperature	Time
1	98°C	00:03:00
2	98°C	00:00:15
3	63°C	00:00:20
4	72°C	00:01:00
5	Go to Step 2, see table	e below for total # of cycles
6	72°C	00:01:00
7	4°C	Hold

Recommended starting point for cycle number optimization.

The optimal number of cycles is a trade-off between generating sufficient final mass for library construction and minimizing PCR amplification artifacts. The number of cDNA cycles should also be reduced if large numbers of cells are sampled.

Targeted Cell Recovery	Total Cycles
<500	13
500–1,000	12



e. Store at 4°C for up to 72 h or –20°C for \leq 1 week, or proceed to the next step.

2.3 cDNA Cleanup – SPRIselect

- **a.** Vortex to resuspend the SPRIselect reagent. Add **60** μ l SPRIselect reagent (**0.6X**) to each sample and pipette mix 15x (pipette set to 150 μ l).
- b. Incubate 5 min at room temperature.
- c. Place on the magnet•High until the solution clears.
- d. Remove the supernatant.
- e. Add 200 μl 80% ethanol to the pellet. Wait 30 sec.
- f. Remove the ethanol.


STOP

- g. Repeat steps e and f for a total of 2 washes.
- h. Centrifuge briefly and place on the magnet•Low.
- i. Remove any remaining ethanol. Air dry for **2 min**. DO NOT exceed **2 min** as this will decrease elution efficiency.
- j. Remove from the magnet. Add 40.5 µl Buffer EB. Pipette mix 15x (pipette set to 35 µl).
- k. Incubate 2 min at room temperature.
- I. Place the tube strip on the magnet•High until the solution clears.
- m.Transfer 40 µl sample to a new tube strip.
- n. Store at 4°C for up to 72 h or at -20°C for up to 4 weeks, or proceed to the next step.

2.4 Post cDNA Amplification QC & Quantification

a. Run **1 μl** sample, diluted 1:10 on an Agilent Bioanalyzer High Sensitivity chip. DO NOT run sample from 2.3B Transferred Supernatant Cleanup step.


For input cells with low RNA content (<1pg total RNA/cell), 1 µl undiluted product may be run. Lower molecular weight product (35 – 150 bp) may be present. This is normal and does not affect sequencing or application performance.




EXAMPLE CALCULATION

i. Select Region

Under the "Electropherogram" view choose the "Region Table". Manually select the region of ~200 - ~9000 bp.



ii. Note Concentration [pg/µl]



iii. <u>Calculate</u>

Multiply the cDNA concentration $[pg/\mu l]$ reported via the Agilent 2100 Expert Software by the elution volume (40 µl) of the Post cDNA Amplification Reaction Clean Up sample (taking any dilution factors into account) and then divide by 1000 to obtain the total cDNA yield in ng.

Example Calculation of cDNA Total Yield

Concentration: 1046.5 pg/µl Elution Volume: 40 Dilution Factor: 10

Total cDNA Yield = Conc'n (pg/µl) x Elution Volume x Dilution Factor

1000 (pg/ng)

= <u>1046.5 (pg/µl) x 40 x 10</u> = 418.6 ng 1000 (pg/ng)

> Carry forward **ONLY 25%** of total cDNA yield into 3' Gene Expression Library Construction (step 3) = 0.25 x Total cDNA yield

> > = 0.25 x 418.6= 104.65 ng

Refer to step 3.5 for appropriate number of Sample Index PCR cycles based on carry forward cDNA yield/input cDNA.

Alternate Quantification Methods See Appendix for representative traces

- Agilent TapeStation
- LabChip

Agilent Bioanalyzer, Agilent TapeStation, LabChip are the recommended methods for accurate quantification.

(If using Qubit Fluorometer and Qubit dsDNA HS Assay Kit, see Appendix)

Step 3

3' Gene Expression Library Construction

- 3.1 Fragmentation, End Repair & A-tailing
- **3.2** Post Fragmentation End Repair & A-tailing Double Sided Size Selection SPRIselect
- 3.3 Adaptor Ligation
- 3.4 Post Ligation Cleanup SPRIselect
- 3.5 Sample Index PCR
- 3.6 Post Sample Index PCR Double Sided Size Selection SPRIselect
- 3.7 Post Library Construction QC

3.0 3' Gene Expression Library Construction




GET STARTED!					
Action	Item	10x PN	Preparation & Handling	Storage	
Room	Fragmentation Buffer	2000091	Vortex, verify no precipitate, centrifuge briefly.	–20°C	
Temperature	🛑 Adaptor Oligos	2000094	Vortex, centrifuge briefly.	–20°C	
	Ligation Buffer	2000092	Vortex, verify no precipitate, centrifuge briefly.	–20°C	
	Dual Index Plate TT Set A Verify name & PN Use indicated plate only	3000431	-	–20°C	
	Beckman Coulter SPRIselect Reagent	-	Manufacturer's recommendations.	-	
	Agilent TapeStation Screen Tape and Reagents If used for QC	-	Manufacturer's recommendations.	-	
	Agilent Bioanalyzer High Sensitivity kit If used for QC	-	Manufacturer's recommendations.	-	
	DNA High Sensitivity Reagent Kit If LabChip used for QC	-	Manufacturer's recommendations.	-	
Place on Ice	Fragmentation Enzyme	2000104	Centrifuge briefly.	–20°C	
	🛑 DNA Ligase	220131	Centrifuge briefly.	–20°C	
(🔿 Amp Mix	2000103	Centrifuge briefly.	–20°C	
	KAPA Library Quantification Kit for Illumina Platforms	-	Manufacturer's recommendations.	-	
Obtain	Qiagen Buffer EB	-	-	Ambient	
	10x Magnetic Separator	230003	See Tips & Best Practices.	Ambient	
	Prepare 80% Ethanol Prepare 10 ml for 4 reactions	-	Prepare fresh.	Ambient	

Step Overview (Step 3.1d)

Correlation between input & library complexity

A Single Cell 3' Gene Expression library is generated using a fixed proportion (10 μ l, 25%) of the total cDNA (40 μ l) obtained at step 2.3m. The complexity of this library will be comparable to one generated using a higher proportion (>25%) of the cDNA. The remaining proportion (30 μ l, 75%) of the cDNA may be stored at 4°C for up to 72 h or at -20°C for longer-term storage (up to 4 weeks).



Note that irrespective of the total cDNA yield (ng), which may vary based on cell type, targeted cell recovery etc., this protocol has been optimized for a broad range of input mass (ng), as shown in the example below. The total number of SI PCR cycles (step 3.5e) should be optimized based on carrying forward a fixed proportion (10 μ l, 25%) of the total cDNA yield calculated during Post cDNA Amplification QC & Quantification (step 2.4).

Cell Type	Targeted	Total cDNA Yield –	cDNA Input into Fragmentation		SI PCR Cycle
Cell Type	Cell Recovery	(ng)	Volume (µl)	Mass (ng)	Number
High RNA Content	Low	250 ng	10 µl	62.5 ng	13
	High O O	1900 ng	10 µl	475 ng	10
Low RNA Content	Low	1 ng	10 µl	0.25 ng	16
	High	200 ng	10 µl	50 ng	12

Example: Library Construction Input Mass & SI PCR Cycles

3.1 Fragmentation, End Repair & A-tailing

a. Prepare a thermal cycler with the following incubation protocol.

	,		
	Lid Temperature	Reaction Volume	Run Time
	65°C	50 µl	~35 min
	Step	Temperature	Time
•	Pre-cool block Pre-cool block prior to preparing the Fragmentation Mix	4°C	Hold
	Fragmentation	32°C	00:05:00
	End Repair & A-tailing	65°C	00:30:00
	Hold	4°C	Hold

- **b.** Vortex Fragmentation Buffer. Verify there is no precipitate.
- c. Prepare Fragmentation Mix on ice. Pipette mix and centrifuge briefly.

Fragmentation Mix <i>Add reagents in the order listed</i>	PN	1Χ (μl)	4X + 10% (µl)	8X + 10% (μl)
Fragmentation Buffer	2000091	5	22	44
Fragmentation Enzyme	2000104	10	44	88
Total	-	15	66	132

d. Transfer ONLY 10 μl purified cDNA sample from Pellet Cleanup (step 2.3m) to a tube strip.

Note that only 10 μ l (25%) cDNA sample transfer is sufficient for generating 3' Gene Expression library.

The remaining **30** µl (75%) cDNA sample can be stored at **4°C** for up to **72** h or at **-20°C** for up to **4 weeks** for generating additional 3' Gene Expression libraries.

- e. Add 25 µl Buffer EB to each sample.
- f. Add 15 µl Fragmentation Mix to each sample.
- g. Pipette mix 15x (pipette set to 35 μ l) on ice. Centrifuge briefly.
- h. Transfer into the pre-cooled thermal cycler (4°C) and press "SKIP" to initiate the protocol.

3.2 Post Fragmentation, End Repair & A-tailing Double Sided Size Selection – SPRIselect

- a. Vortex to resuspend SPRIselect reagent. Add **30 µl** SPRIselect **(0.6X)** reagent to each sample. Pipette mix 15x (pipette set to 75 µl).
- b. Incubate 5 min at room temperature.
- c. Place on the magnet•High until the solution clears. DO NOT discard supernatant.



- d. Transfer **75** µl supernatant to a new tube strip.
- e. Vortex to resuspend SPRIselect reagent. Add 10 μl SPRIselect reagent (0.8X) to each sample. Pipette mix 15x (pipette set to 80 μl).
- f. Incubate 5 min at room temperature.
- g. Place on the magnet•High until the solution clears.





- h. Remove 80 µl supernatant. DO NOT discard any beads.
- i. Add 125 µl 80% ethanol to the pellet. Wait 30 sec.
- j. Remove the ethanol.
- k. Repeat steps i and j for a total of 2 washes.
- **I.** Centrifuge briefly. Place on the magnet•Low until the solution clears. Remove remaining ethanol. DO NOT over dry to ensure maximum elution efficiency.
- **m.** Remove from the magnet. Add **50.5 μl** Buffer EB to each sample. Pipette mix 15x (pipette set to 45 μl).
- n. Incubate 2 min at room temperature.
- o. Place on the magnet•High until the solution clears.
- p. Transfer 50 µl sample to a new tube strip.

3.3 Adaptor Ligation

a. Prepare Adaptor Ligation Mix. Pipette mix and centrifuge briefly.

Adaptor Ligation Mix Add reagents in the order listed	PN	1X (μl)	4X + 10% (µl)	8X + 10% (μl)
Ligation Buffer	2000092	20	88	176
DNA Ligase	220131	10	44	88
Adaptor Oligos	2000094	20	88	176
Total	-	50	220	440

b. Add **50 \mul** Adaptor Ligation Mix to **50 \mul** sample. Pipette mix 15x (pipette set to 90 μ l). Centrifuge briefly.

c. Incubate in a thermal cycler with the following protocol.

l

Lid Temperature	Reaction Volume	Run Time
30°C	100 µl	15 min
Step	Temperature	Time
1	20°C	00:15:00
2	4°C	Hold

3.4 Post Ligation Cleanup – SPRIselect

- **a.** Vortex to resuspend SPRIselect Reagent. Add **80 μl** SPRIselect Reagent **(0.8X)** to each sample. Pipette mix 15x (pipette set to 150 μl).
- b. Incubate 5 min at room temperature.
- c. Place on the magnet•High until the solution clears.
- d. Remove the supernatant.
- e. Add 200 μl 80% ethanol to the pellet. Wait 30 sec.
- f. Remove the ethanol.
- g. Repeat steps e and f for a total of 2 washes.
- h. Centrifuge briefly. Place on the magnet•Low.
- Remove any remaining ethanol. Air dry for 2 min.
 DO NOT exceed 2 min as this will decrease elution efficiency.
- j. Remove from the magnet. Add 30.5 µl Buffer EB. Pipette mix 15x.
- k. Incubate 2 min at room temperature.
- I. Place on the magnet•Low until the solution clears.
- m. Transfer 30 µl sample to a new tube strip.

3.5 Sample Index PCR

a. Choose the appropriate sample index sets to ensure that no sample indices overlap in a multiplexed sequencing run. Record the 10x Sample Index name (PN-3000431 Dual Index Plate TT Set A well ID) used.

- **b.** Add **50 μl** Amp Mix (2000103) to **30 μl** sample.
- c. Add 20 µl of an individual Dual Index TT Set A to each sample and record the well ID used. Pipette mix 5x (pipette set to 90 µl). Centrifuge briefly.
- d. Incubate in a thermal cycler with the following protocol.

Reaction Volume	Run Time
100 µl	~25-40 min
Temperature	Time
98°C	00:00:45
98°C	00:00:20
54°C	00:00:30
72°C	00:00:20
Go to step 2, see bel	ow for # of cycles
72°C	00:01:00
4°C	Hold
	100 μl Temperature 98°C 98°C 54°C 52°C Go to step 2, see bel 72°C



The total cycles should be optimized based on 25% carry forward cDNA yield/input calculated during cDNA QC & Quantification (step 2.4)

cDNA Input	Total Cycles
0.25-25 ng	14-16
25-150 ng	12-14
150-500 ng	10-12
500-1,000 ng	8-10
1,000-1,500 ng	6-8
>1500 ng	5

STOP

f. Store at 4°C for up to 72 h or proceed to the next step.

DUAL


INDE>

3.6 Post Sample Index PCR Double Sided Size Selection – SPRIselect

- a. Vortex to resuspend the SPRIselect reagent. Add **60 µl** SPRIselect Reagent **(0.6X)** to each sample. Pipette mix 15x (pipette set to 150 µl).
- b. Incubate 5 min at room temperature.
- c. Place the magnet•High until the solution clears. DO NOT discard supernatant.
- d. Transfer 150 µl supernatant to a new tube strip.
- e. Vortex to resuspend the SPRIselect reagent. Add **20 µl** SPRIselect Reagent **(0.8X)** to each sample. Pipette mix 15x (pipette set to 150 µl).
- f. Incubate 5 min at room temperature.
- g. Place the magnet•High until the solution clears.
- h. Remove 165 µl supernatant. DO NOT discard any beads.
- i. With the tube still in the magnet, add $200 \,\mu l \, 80\%$ ethanol to the pellet. Wait $30 \, sec.$
- j. Remove the ethanol.
- k. Repeat steps i and j for a total of 2 washes.
- I. Centrifuge briefly. Place on the magnet•Low. Remove remaining ethanol.
- **m.** Remove from the magnet. Add **35.5 μl** Buffer EB. Pipette mix 15x.
- n. Incubate 2 min at room temperature.
- o. Place on the magnet•Low until the solution clears.
- **p.** Transfer **35** μl to a new tube strip.
- q. Store at 4°C for up to 72 h or at -20°C for long-term storage.

3.7 Post Library Construction QC

Run 1 µl sample at 1:10 dilution on an Agilent Bioanalyzer High Sensitivity chip.



Determine the average fragment size from the Bioanalyzer trace. This will be used as the insert size for library quantification.

Alternate QC Method:

- Agilent TapeStation
- LabChip

See Appendix for representative traces

See Appendix for Post Library Construction Quantification

Step 3

Sequencing

Sequencing Libraries

Single Cell 3' Gene Expression libraries comprise standard Illumina paired-end constructs which begin with P5 and end with P7. 16 bp 10x Barcodes are encoded at the start of TruSeq Read 1, while 10 bp i5 and i7 sample index sequences are incorporated as sample index reads. TruSeq Read 1 and Read 2 are standard Illumina sequencing primer sites used in paired-end sequencing. TruSeq Read 1 is used to sequence 16 bp 10x Barcodes and 12 bp UMI. Sequencing these libraries produces a standard Illumina BCL data output folder.

Chromium Single Cell 3' Gene Expression Dual Index Library



Illumina Sequencer Compatibility

The compatibility of the listed sequencers has been verified by 10x Genomics. Some variation in assay performance is expected based on sequencer choice. For more information about performance variation, visit the 10x Genomics Support website.

- MiSeq
- NextSeq 500/550
- NextSeq 1000/2000
- HiSeq 2500 (Rapid Run)
- HiSeg 3000/4000
- NovaSeq

Sample Indices

Each sample index in the Dual Index Kit TT Set A (PN-1000215) is a mix of one unique i7 and one unique i5 sample index. If multiple samples are pooled in a sequence lane, the sample index name (i.e. the Dual Index TT Set A plate well ID, SI-TT-) is needed in the sample sheet used for generating FASTQs with "cellranger mkfastq". Samples utilizing the same sample index should not be pooled together, or run on the same flow cell lane, as this would not enable correct sample demultiplexing.

3' Gene Expression Library Sequencing Depth & Run Parameters

Sequencing Depth	Minimum 20,000 read pairs per cell
Sequencing Type	Paired-end, dual indexing
Sequencing Read	Recommended Number of Cycles
Read 1 i7 Index i5 Index Read 2	28 cycles 10 cycles 10 cycles 90 cycles

Library Loading

Once quantified and normalized, the 3' Gene Expression libraries should be denatured and diluted as recommended for Illumina sequencing platforms. Refer to Illumina documentation for denaturing and diluting libraries. Refer to the 10x Genomics Support website, for more information.

Instrument	Loading Concentration (pM)	PhiX (%)
MiSeq	11	1
NextSeq 500/550	1.8	1
NextSeq 1000/2000	650	1
HiSeq 2500 (RR)	11	1
HiSeq 4000	240	1
NovaSeq	150*/300	1

* Use 150 pM loading concentration for Illumina XP workflow.

Library Pooling

The 3' Gene Expression libraries may be pooled for sequencing, taking into account the differences in cell number and per-cell read depth requirements between each library. Samples utilizing the same sample index should not be pooled together, or run on the same flow cell lane, as this would not enable correct sample demultiplexing.

Data Analysis and Visualization

Sequencing data may be analyzed using Cell Ranger or 10x Genomics Cloud Analysis and visualized using Loupe Browser. Key features for these tools are listed below. For detailed productspecific information, visit the 10x Genomics Support website.

Cell Ranger

Cell Ranger is a set of analysis pipelines that processes Chromium Single Gene Expression data to align reads, generate Feature Barcode matrices and perform clustering and gene expression analysis.

- Input: Base call (BCL) and FASTQ
- Output: BAM, MEX, CSV, HDF5, Web Summary, .cloupe/.loupe
- Operating System: Linux

Cloud Analysis

Cloud Analysis is currently only available for US customers.

Cloud Analysis allows users to run Cell Ranger analysis pipelines from a web browser while computation is handled in the cloud.

- Key features: scalable, highly secure, simple to set up and run
- Input: FASTQ
- Output: BAM, MEX, CSV, HDF5, Web Summary, .cloupe/.loupe.

Loupe Browser

Loupe Browser is an interactive data visualization tool that requires no prior programming knowledge.

- Input: .cloupe
- Output: Data visualization, including t-SNE and UMAP projections, custom clusters, differentially expressed genes
- Operating System: MacOS, Windows







Troubleshooting

GEMs

STEP NORMAL

1.4 d After Chip L is removed from the Controller and the wells are exposed



All 8 recovery wells are similar in volume and opacity.

REAGENT CLOGS & WETTING FAILURES




Recovery well G indicates a reagent clog. Recovery well C and E indicate a wetting failure. Recovery wells B, D, and F are normal. Wells A and H contain 50% Glycerol Solution.

1.4 f Transfer GEMs from Chip L Row Labeled 3



All liquid levels are similar in volume and opacity without air trapped in the pipette tips.



Pipette tips C and E indicate a wetting failure. Pipette tip C contains partially emulsified GEMs. Emulsion is absent in pipette tip E. Pipette tip G indicates a reagent clog.

2.1 a


GEMs +

NORMAL



All liquid levels are similar in the aqueous sample volume (clear) and Recovery Agent/Partitioning Oil (pink).

REAGENT CLOGS & WETTING FAILURES



Tube G indicates a reagent clog has occurred. There is a decreased volume of aqueous layer (clear). Tube C and E indicate a wetting failure has


occurred. There is an abnormal volume of Recovery Agent/Partitioning Oil (pink).

2.1 b After aspiration of Recovery Agent/ Partitioning Oil


STEP

After transfer of the

Recovery Agent



All liquid volumes are similar in the aqueous sample volume (clear) and residual Recovery Agent/Partitioning Oil (pink).



Tube G indicates a reagent clog has occurred. There is a decreased volume of aqueous layer (clear). There is also a greater residual volume of Recovery Agent/Partitioning Oil (pink). Tube C and E indicate a wetting failure has occurred. There is an abnormal residual volume of Recovery Agent/Partitioning Oil (pink).

2.1 d After addition of Dynabeads Cleanup Mix



All liquid volumes are similar after addition of the Dynabeads Cleanup Mix.





Tube G indicates a reagent clog has occurred. There is an abnormal ratio of Dynabeads Cleanup Mix (brown) to Recovery Agent/Partitioning Oil (appears white).

Tube C and E indicate a wetting failure has occurred. There is an abnormal ratio of Dynabeads Cleanup Mix (brown) to Recovery Agent/Partitioning Oil (appears white).

Chromium Controller Errors

If the Chromium Controller or the Chromium Single Cell Controller fails to start, an error tone will sound and one of the following error messages will be displayed:

- a. Chip not read Try again: Eject the tray, remove and/or reposition the Chromium Next GEM Secondary Holder assembly and try again. If the error message is still received after trying this more than twice, contact support@10xgenomics.com for further assistance.
- **b.** Check gasket: Eject the tray by pressing the eject button to check that the 10x Gasket is correctly installed on the Chromium Next GEM Chip. If the error message persists, contact support@10xgenomics.com for further assistance.
- c. Error Detected: Row _ Pressure:
 - i. If this message is received within a few seconds of starting a run, eject the tray by pressing the eject button and check for dirt or deposits on the 10x Gasket. If dirt is observed, replace with a new 10x Gasket and try again. If the error message is still received after trying this more than twice, contact support@10xgenomics.com for further assistance.
 - ii. If this message is received after a few minutes into the run, the Chromium Next GEM Chip must be discarded. **Do not try running this Chromium Next GEM Chip** again as this may damage the Chromium Controller.
- d. Invalid Chip CRC Value: This indicates that a Chromium Next GEM Chip has been used with an older firmware version. The chip must be discarded. Contact support@10xgenomics.com for further assistance.
- e. Chip Holder Not Present: Open the controller drawer and check if chip holder is present. Insert chip properly into chip holder and retry.
- f. Unauthorized Chip: This indicates that an incompatible non-Next GEM chip has been used with an instrument that only can run Next GEM assays. Use only Chromium Controller (PN-120223;120246) or Chromium Single Cell Controller (PN-120263;120212) to run that chip or chip must be discarded. Contact support@10xgenomics.com for further assistance.
- **g. Endpoint Reached Early:** If this message is received, contact support@10xgenomics.com for further assistance.

Appendix

Post Library Construction Quantification Agilent TapeStation Traces LabChip Traces Oligonucleotide Sequences

Post Library Construction Quantification

- a. Thaw KAPA Library Quantification Kit for Illumina Platforms.
- b. Dilute 2 µl sample with deionized water to appropriate dilutions that fall within the linear detection range of the KAPA Library Quantification Kit for Illumina Platforms. (For more accurate quantification, make the dilution(s) in duplicate).
- **c.** Make enough Quantification Master Mix for the DNA dilutions per sample and the DNA Standards (plus 10% excess) using the guidance for 1 reaction volume below.

Quantification Master Mix	1X (μl)
SYBR Fast Master Mix + Primer	12
Water	4
Total	16

- **d.** Dispense **16** µl Quantification Master Mix for sample dilutions and DNA Standards into a 96 well PCR plate.
- e. Add 4 µl sample dilutions and 4 µl DNA Standards to appropriate wells. Centrifuge briefly.
- f. Prepare a qPCR system with the following protocol. Insert the plate and start the program.

Step	Temperature	Run Time
1	95°C	00:03:00
2	95°C	00:00:05
3	67°C	00:00:30
4	Go to Step 2, 29X (Total 30 cycles)	


g. Follow the manufacturer's recommendations for qPCR-based quantification. For library quantification for sequencer clustering, determine the concentration based on insert size derived from the Bioanalyzer/TapeStation trace.

Agilent TapeStation Traces

Agilent TapeStation Traces


Agilent TapeStation High Sensitivity D5000 ScreenTape was used. Protocol steps correspond to the Chromium Next GEM Single Cell 3' LT (Dual Index) v3.1 User Guide (CG000399)

Protocol Step 2.4 – cDNA QC & Quantification



Run 2 μl sample mixed with 2 μl loading buffer. Ensure dilution factor is factored in when calculating cDNA yield/µl.

Protocol Step 3.7 – Post Library Construction QC



Run 2 µl diluted sample (1:10 dilution) mixed with 2 µl loading buffer.


All traces are representative.

LabChip Traces


LabChip Traces

DNA High Sensitivity Reagent Kit was used. Protocol steps correspond to the Chromium Next GEM Single Cell 3' LT Reagent Kits v3.1 (Dual Index) User Guide (CG000399)

Protocol Step 2.4 - cDNA QC & Quantification



Protocol Step 3.7 – Post Library Construction QC




Alternate QC Method:

Qubit Fluorometer and Qubit dsDNA HS Assay Kit

Multiply the cDNA concentration reported via the Qubit Fluorometer by the elution volume (40 µl) to obtain the total cDNA yield in ng. To determine the equivalent range using the Agilent 2100 Expert Software, select the region encompassing 35-10,000 bp.

Oligonucleotide Sequences



Oligonucleotide Sequences

