



Updated calibration model for ALPHA II

From today 29/06/2023 this is the only valid method to use, and all further analysis shall be done using this new model, until anything is announced.

(FIS_IBU Fluor_finder vs230629.q2 file)