



Summary of Safety and Clinical Performance
for
Precipitating Hydrophobic Injectable Liquid
PHIL™ System
SSCP23-0016

MicroVention, Inc.
35 Enterprise
Aliso Viejo, California 92656, USA

MicroVention Europe SARL
30 bis, rue du Vieil Abrevoir
78100 Saint-Germain-en-Laye, France

DOCUMENT CHANGE HISTORY

SSCP Revision	Change Description	NB approved/verified
A	Initial release	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No* Validation language:

*Annual entries must be included. An entry stating such must be added if a revision is not required.

TABLE OF CONTENTS

1	SUMMARY OF SAFETY AND CLINICAL PERFORMANCE	4
1.1	Device Identification and General Information	4
1.2	Intended Purpose of the Device	4
1.3	Device Description.....	5
1.4	Risks and Warnings	8
1.4.1	Residual Risks and Undesirable Effects	8
1.4.2	Warnings and Precautions.....	8
1.4.3	Potential Complications / Adverse Effects	11
1.4.4	Other Aspects of Safety	12
1.5	Summary of the Clinical Evaluation and PMCF	12
1.5.1	Equivalent Device Clinical Data.....	12
1.5.2	Pre-CE-Mark Clinical Data.....	13
1.5.3	Clinical Data	13
1.5.4	Clinical Performance and Safety	14
1.5.5	Post-Market Clinical Follow-up	15
1.6	Possible Diagnostic or Therapeutic Alternatives	16
1.6.1	Treatment Options and Interventions.....	16
1.6.2	Available Technologies	18
1.7	Suggested Profile and Training for Users.....	19
1.8	Reference to any Harmonized Standards and CS	19
1.9	References.....	23

LIST OF TABLES

Table 1.1	Device Identification and General Information	4
Table 1.2	Intended Use	4
Table 1.3	Device Description	5
Table 15:	Benefits/Risk of Standard of Care	Error! Bookmark not defined.
Table 16:	Similar Devices.....	18
Table 17:	Similar Device Benefits/Risks	Error! Bookmark not defined.

1 SUMMARY OF SAFETY AND CLINICAL PERFORMANCE

This Summary of Safety and Clinical Performance (SSCP) is intended to provide public access to an updated summary of the main aspects of the safety and clinical performance of the device.

The SSCP is not intended to replace the Instructions For Use (IFU) as the main document to ensure the safe use of the device, nor is it intended to provide diagnostic or therapeutic suggestions to intended users or patients.

The following information is intended for users/healthcare professionals.

1.1 Device Identification and General Information

Table 1-1 Device Identification and General Information

Device Names	
Device Trade Name	Precipitating Hydrophobic Injectable Liquid (PHIL) Liquid Embolic System
EMDN Code	C010402020302 Embolization Fluids
Medical Device Nomenclature (EMDN)	60939- Neurovascular embolization plug
Device Class	III
Basic UDI-DI	37015174PHILKC
Year when first certificate (CE) was issued for the device	2014
Legal Manufacturer	
Name & Address	MicroVention Europe SARL (MVE) 30 bis, rue du Vieil Abrevoir 78100 Saint Germain-en-Laye, France
Manufacturer SRN	MVE: FR-MF-000004449
Notified Body	
Name & Address	DQS Medizinprodukte GmbH August-Schanz-Straße 21 D-60433 Frankfurt am Main Germany
Notified Body Identification Number	0297

1.2 Intended Purpose of the Device

Table 1-2 Intended Use

Intended Purpose

Intended Purpose	The PHIL device is intended for use in the embolization of lesions in the peripheral and neurovasculature, including arteriovenous malformations and hypervascular tumors.
Indications for Use	The PHIL device is intended for use in the embolization of lesions in the peripheral and neurovasculature, including arteriovenous malformations and hypervascular tumors.
Target Population	Endovascular embolization using the PHIL device can be used in a wide range of procedures for patients requiring embolization of neurovascular or peripheral vascular lesions such as brain arteriovenous malformations (bAVMs), dural arteriovenous fistulas (dAVFs), peripheral AVMs (e.g., pulmonary AVM), peripheral AVFs, pseudoaneurysms, and hypervascular tumors.
Contraindications and/or Limitations	The use of the PHIL device is contraindicated when any of the following conditions exist: <ul style="list-style-type: none"> • When patient has severe iodine allergy. • When optimal microcatheter placement is not possible. • When provocative testing indicates intolerance to the occlusion procedure. • When vasospasm stops blood flow. • Not for use with premature infants (<1,500 g) or individuals with significant liver and kidney function impairment.

1.3 Device Description

Table 1-3 Device Description

Device Description																	
Description of the Device	<p>The PHIL device is a non-adhesive liquid embolic agent comprised of a co-polymer dissolved in DMSO (dimethyl sulfoxide). An iodine component is chemically bonded to the co-polymer to provide a radiopaque element during fluoroscopic visualization. A PHIL Liquid Embolic Starter Kit System consists of a sterile, pre-filled, 1 mL syringe of PHIL liquid embolic, a sterile, pre-filled 1 mL syringe of DMSO, and a Universal microcatheter Adaptor. A Re-PHIL Liquid Embolic System consists of two sterile, pre-filled 1 mL syringes of PHIL liquid embolic and a Universal microcatheter Adaptor. A DMSO compatible delivery microcatheter that is indicated for use in the neurovascular or peripheral vasculature is used to access the embolization target site. The PHIL Liquid Embolic System is available in several product formulations: PHIL 35%, PHIL 30%, PHIL 25%, and low viscosity PHIL LV liquid embolics. PHIL LV and PHIL 25% liquid embolic devices travel more distally and penetrate deeper into the nidus due to their lower viscosities compared to PHIL 30% or 35% liquid embolic devices. It is recommended that PHIL LV liquid embolic is used with a flow-arrest technique, such as a balloon catheter, to increase injection control. Final solidification of PHIL material occurs within three minutes for all product formulations.</p> <table border="1" style="margin-left: auto; margin-right: auto;"> <thead> <tr> <th>Catalog Number</th> <th>Concentration</th> </tr> </thead> <tbody> <tr> <td>LEN10250</td> <td>PHIL 25%</td> </tr> <tr> <td>LEN10250RE</td> <td>REPHIL 25% (Refill Kit)</td> </tr> <tr> <td>LEN10300</td> <td>PHIL 30%</td> </tr> <tr> <td>LEN10300RE</td> <td>REPHIL 30% (Refill Kit)</td> </tr> <tr> <td>LEN10350</td> <td>PHIL 35%</td> </tr> <tr> <td>LEN10350RE</td> <td>REPHIL 35% (Refill Kit)</td> </tr> <tr> <td>LEN10LV250</td> <td>PHIL LV</td> </tr> </tbody> </table>	Catalog Number	Concentration	LEN10250	PHIL 25%	LEN10250RE	REPHIL 25% (Refill Kit)	LEN10300	PHIL 30%	LEN10300RE	REPHIL 30% (Refill Kit)	LEN10350	PHIL 35%	LEN10350RE	REPHIL 35% (Refill Kit)	LEN10LV250	PHIL LV
Catalog Number	Concentration																
LEN10250	PHIL 25%																
LEN10250RE	REPHIL 25% (Refill Kit)																
LEN10300	PHIL 30%																
LEN10300RE	REPHIL 30% (Refill Kit)																
LEN10350	PHIL 35%																
LEN10350RE	REPHIL 35% (Refill Kit)																
LEN10LV250	PHIL LV																

	LEN10LV250RE	REPHIL LV (Refill Kit)																								
Design Characteristics of the Device	<p>The PHIL liquid embolic is a non-adhesive co-polymer comprised of triiodophenol-(lactide-co-glycolide) acrylate and hydroxyethyl methacrylate. An iodine component is chemically bonded to the co-polymer to provide a radiopacifier element during fluoroscopic visualization. The PHIL material is delivered in liquid phase through a microcatheter to the target lesion under fluoroscopic control. The mode of action of the copolymer is its solubility in the organic solvent (DMSO), but insolubility in an aqueous environment. Upon contact with blood, the liquid embolic precipitates in-situ resulting in a polymeric embolus localized in the vasculature. This occlusion process of the target lesion is dependent on many variables such as delivery rate of embolic, local rate of blood flow, and vascular dimensions. These factors will determine the direction of movement and penetration distance of the liquid embolic as it rapidly solidifies to form an outer crust and maintains a liquid center to promote distal propagation. PHIL Embolic System is an implantable device with direct intravascular tissue and blood contact for long-term patient contact duration.</p> <table border="1"> <thead> <tr> <th>Key Functional Element</th> <th>Materials</th> </tr> </thead> <tbody> <tr> <td>Flushing Solution</td> <td>Dimethyl sulfoxide (DMSO)</td> </tr> <tr> <td>PHIL Embolic Liquid</td> <td>Tri-iodophenol-(lactide-coglycolide) acrylate and hydroxyethyl methacrylate dissolved in DMSO</td> </tr> <tr> <td rowspan="5">Syringe</td> <td>Barrel: Borosilicate glass</td> </tr> <tr> <td>Luer: Stainless Steel 316/316L</td> </tr> <tr> <td>Stopper: Ethylene propylene diene monomer</td> </tr> <tr> <td>Plunger: Stainless Steel 304</td> </tr> <tr> <td>Cap: Polypropylene</td> </tr> <tr> <td rowspan="4">Universal Adaptor</td> <td>Finger grip: Polypropylene (yellow or white)</td> </tr> <tr> <td>Hub: Polypropylene, 304/304L Stainless Steel</td> </tr> <tr> <td>Tuohy Borst Body: Polypropylene</td> </tr> <tr> <td>Tuohy Borst Cap: Polypropylene</td> </tr> <tr> <td rowspan="4">Tray Packaging</td> <td>Tuohy Gasket: Silicone</td> </tr> <tr> <td>Tray: Polycarbonate</td> </tr> <tr> <td>Tray Lid: Tyvek</td> </tr> <tr> <td>Insert: Solid Bleached Sulfate Paperboard</td> </tr> <tr> <td></td> <td>Carton: Solid Bleached Sulfate Paperboard</td> </tr> </tbody> </table>		Key Functional Element	Materials	Flushing Solution	Dimethyl sulfoxide (DMSO)	PHIL Embolic Liquid	Tri-iodophenol-(lactide-coglycolide) acrylate and hydroxyethyl methacrylate dissolved in DMSO	Syringe	Barrel: Borosilicate glass	Luer: Stainless Steel 316/316L	Stopper: Ethylene propylene diene monomer	Plunger: Stainless Steel 304	Cap: Polypropylene	Universal Adaptor	Finger grip: Polypropylene (yellow or white)	Hub: Polypropylene, 304/304L Stainless Steel	Tuohy Borst Body: Polypropylene	Tuohy Borst Cap: Polypropylene	Tray Packaging	Tuohy Gasket: Silicone	Tray: Polycarbonate	Tray Lid: Tyvek	Insert: Solid Bleached Sulfate Paperboard		Carton: Solid Bleached Sulfate Paperboard
Key Functional Element	Materials																									
Flushing Solution	Dimethyl sulfoxide (DMSO)																									
PHIL Embolic Liquid	Tri-iodophenol-(lactide-coglycolide) acrylate and hydroxyethyl methacrylate dissolved in DMSO																									
Syringe	Barrel: Borosilicate glass																									
	Luer: Stainless Steel 316/316L																									
	Stopper: Ethylene propylene diene monomer																									
	Plunger: Stainless Steel 304																									
	Cap: Polypropylene																									
Universal Adaptor	Finger grip: Polypropylene (yellow or white)																									
	Hub: Polypropylene, 304/304L Stainless Steel																									
	Tuohy Borst Body: Polypropylene																									
	Tuohy Borst Cap: Polypropylene																									
Tray Packaging	Tuohy Gasket: Silicone																									
	Tray: Polycarbonate																									
	Tray Lid: Tyvek																									
	Insert: Solid Bleached Sulfate Paperboard																									
	Carton: Solid Bleached Sulfate Paperboard																									

	The PHIL Embolic System does not incorporate a medicinal substance, animal tissues or blood products.														
Previous Generations or Variants, if applicable	<table border="1"> <thead> <tr> <th data-bbox="496 359 672 464"></th> <th data-bbox="672 359 812 464">Model Name</th> <th data-bbox="812 359 1192 464">Characteristics</th> <th data-bbox="1192 359 1382 464">Status</th> </tr> </thead> <tbody> <tr> <td data-bbox="496 464 672 814">Original Device</td> <td data-bbox="672 464 812 814"> PHIL 25% PHIL 30% PHIL 35% </td> <td data-bbox="812 464 1192 814"> <u>PHIL 25%</u> Low flow scenario, Distal access <u>PHIL 30%</u> Moderate flow scenarios, When feeding pedicle injection conducted close to the nidus/ fistula <u>PHIL 35%</u> Higher flow scenarios, Larger component embolization </td> <td data-bbox="1192 464 1382 915" rowspan="2">Currently Manufactured</td> </tr> <tr> <td data-bbox="496 814 672 915">Incremental Development</td> <td data-bbox="672 814 812 915">PHIL LV</td> <td data-bbox="812 814 1192 915">Low flow scenarios, Distal access</td> </tr> </tbody> </table>		Model Name	Characteristics	Status	Original Device	PHIL 25% PHIL 30% PHIL 35%	<u>PHIL 25%</u> Low flow scenario, Distal access <u>PHIL 30%</u> Moderate flow scenarios, When feeding pedicle injection conducted close to the nidus/ fistula <u>PHIL 35%</u> Higher flow scenarios, Larger component embolization	Currently Manufactured	Incremental Development	PHIL LV	Low flow scenarios, Distal access			
	Model Name	Characteristics	Status												
Original Device	PHIL 25% PHIL 30% PHIL 35%	<u>PHIL 25%</u> Low flow scenario, Distal access <u>PHIL 30%</u> Moderate flow scenarios, When feeding pedicle injection conducted close to the nidus/ fistula <u>PHIL 35%</u> Higher flow scenarios, Larger component embolization	Currently Manufactured												
Incremental Development	PHIL LV	Low flow scenarios, Distal access													
Single use – sterilization method	The PHIL device is sterilized by steam sterilization and is intended for single use only. The PHIL device kit is intended to be used for one patient. Do not resterilize and/or reuse the device. Reuse and/or resterilization can increase risk of infection, cause a pyrogenic response or other life- threatening complications.														
Description of Accessories	<ol style="list-style-type: none"> DMSO: PHIL co-polymer is required to be dissolved in DMSO (Dimethyl Sulfoxide) with an iodine component that is chemically bonded to the co-polymer to provide fluoroscopic visualization. The PHIL device is delivered in liquid phase through a DMSO primed microcatheter placed at the target lesion under fluoroscopic control. The universal adaptor: The universal adaptor connects the syringe and microcatheters, with the tip of the adaptor tube positioned within the hub inner diameter to minimize mixing (dilution with DMSO) of the Liquid Embolic solution as it leaves the syringe and is delivered through the microcatheter. 														
Description of other Devices or Products intended to be used in combination	<ol style="list-style-type: none"> Platinum coils may be used to effectively slow the flow, thus facilitating the following embolization with a liquid embolic agent. Microcatheters/Catheters, especially the recently developed detachable tip microcatheters/catheters have facilitated the removal of the catheter from hardened liquid embolic casts. Balloon microcatheters/catheters are used to enhance the protection of normal vasculature and improve the penetration of liquid embolics into large bAVMs/AVMs, which led to a reduction in the procedure times and radiation exposure, in particular, dual-lumen balloon microcatheter (such as the Scepter Mini) may allow a more efficient and controlled injection of liquid embolic agents in smaller, more distal arteries (>1.7 mm), thereby reducing risks of reflux. 														

1.4 Risks and Warnings

1.4.1 Residual Risks and Undesirable Effects

Hazards associated with the use of the PHIL Liquid Embolic System are assessed and risks of the resulting harms are minimized through the use of risk mitigation/control measures. All known foreseeable risks have been evaluated and mitigated.

Risks associated with the PHIL Liquid Embolic System include the following:

- Device not biocompatible or not sterile; packaging failure
- Dimethyl Sulfoxide (DMSO) introduce too quickly
- Liquid embolic migrate outside of targeted area due to misuse of product
- Device used after expiration date
- Device used outside of the intended use / Off label use of device -Exposure to liquid embolic material for unintended application and/or population (patient), which results in catastrophic health effects, may lead to death.
- Unable to inject Liquid Embolic material to operate intended area due to Liquid Embolic material occlusion in catheter lumen
- Unable to inject Liquid Embolic material into microcatheter due to device contamination from waste products and or medical device disposal
- Catheter breakage or rupture due to improper use of catheter, excessive pressure to inject liquid embolic material, or incompatibility of catheter
- Product is an irritant to vessel and/or patient intolerability
- Improper anticoagulation or catheter degradation during procedure
- Lack of cohesive/consistency of product or foreign body embolism due to incorrect formulation
- Catheter entrapment or difficult to remove catheter due to vasospasm; product reflux or very distal afferent; lengthened and tortuous pedicle
- Inadequate visibility under fluoroscopy or PHIL device deployment incorrectly due to incorrect formulation or material selection
- PHIL liquid change in brown / dark color due to expose to higher temperature than suggested

1.4.2 Warnings and Precautions

The warnings and precautions/cautions for the PHIL device are:

Warnings

- Performing embolization to occlude blood vessels is a high-risk procedure. This device should be used only by physicians with neuro or peripheral interventional training and a thorough knowledge of the vascular pathology to be treated, vascular architecture, angiographic techniques, and super-selective embolization techniques.

- Vascular malformation embolization may influence or change blood flow patterns, thereby subjecting arteries supplying the vascular malformation or the normal surrounding tissues and perivascular space around the malformation to increased pressures, or by causing an increase in the intra-nidal pressure. These conditions could result in hemorrhagic complications.
- Care must be taken to avoid venous outflow occlusion prior to nidal or arterial feeders as this may also produce malformation hemorrhage. If the PHIL device extravasates outside the vascular space, secondary to vessel wall compromise, a sub-acute inflammatory response to the material may occur in the surrounding vascular space which may lead to tissue damage.
- There may be some topical hypersensitivity and/or release of histamines from Dimethyl Sulfoxide (DMSO).
- Therapeutic embolization should not be performed when high blood flow precludes safe delivery of the embolic agent to prevent non-target embolization.
- **The microcatheter tip should be placed as distal as possible, and as close to the target vascular lesion as possible to prevent any non-target embolization of normal surrounding tissue or cranial nerves.**
- Premature solidification of the PHIL device may occur if the microcatheter or luer hub comes in contact with any solution such as saline, blood or contrast.
- Use only DMSO compatible microcatheters indicated for use in the neuro or peripheral vasculature. Other microcatheters or syringes may not be compatible with DMSO and their use can result in thromboembolic events due to microcatheter degradation.
- Use only the MicroVention pre-filled syringes to inject DMSO and the PHIL device. Other syringes may not be compatible with DMSO.
- Rapid injection of DMSO into the vasculature space may lead to vasospasm and/or angionecrosis.
- In the event of microcatheter occlusion, excessive force to advance the plunger may result in microcatheter rupture due to over pressurization.
- Do not allow more than **1 cm of the PHIL device** to reflux back over microcatheter tip. Excessive reflux may result in difficult microcatheter removal.
- After using a microcatheter with the PHIL device, do not attempt to clear or inject any material through it. Such attempts may lead to embolus or non-target embolization.
- STOP injection if the PHIL device is not visualized exiting the microcatheter tip. If the microcatheter becomes occluded, over-pressurization and vessel rupture can occur. During

injection, continuously verify under fluoroscopy that the PHIL device is exiting the microcatheter tip.

- STOP injection if increased resistance is observed. If increased resistance occurs, determine the cause (e.g. occlusion in microcatheter lumen) and replace the microcatheter if needed. Do not attempt to clear or overcome resistance by applying increased injection pressure, as use of excessive pressure may result in microcatheter or vessel rupture and embolization of non-target areas.
- Only use thumb pressure to inject the PHIL device. Using the palm of hand to advance plunger may result in microcatheter or vessel rupture due to over pressurization in the event of microcatheter occlusion.
- DO NOT interrupt the PHIL device injection for longer than three minutes prior to re-establishing injection. This may cause solidification of the PHIL device within the microcatheter tip resulting in microcatheter occlusion. Use of excessive pressure to clear the microcatheter may result in microcatheter or vessel rupture and embolization of non-target areas.
- The PHIL Liquid Embolic System is provided sterile and non-pyrogenic unless the unit package is opened or damaged. Do not use if the packaging damaged. Use before expiration date noted on the product packaging.
- The PHIL Liquid Embolic System is intended for single use only. The kit is intended to be used for one patient. Do not resterilize and/or reuse the device. Reuse and/or resterilization can increase risk of injection, cause a pyrogenic response or other life-threatening complications.
- Caution is to be used for nickel-sensitive and/or pregnant female patients, as the device packaged in glass syringes may contain Nickel.

Cautions

- Do not use if packaging is opened or damaged.
- This device is intended for single use only. Do not reuse, reprocess, or re-sterilize.
- Reuse, reprocessing, or re-sterilization may compromise the integrity of the device and /or lead to device failure which, in turn, may result in patient injury, illness, or death. Reuse, reprocessing, or re-sterilization may also create a risk of contamination of the device and/or cause patient infection or cross-infection, including but not limited to, the transmission of infectious disease(s) from one patient to another. Contamination of the device may lead to injury, illness or death of the patient.
- After use, dispose in accordance with hospital, administrative and/or local government policy.

Precautions

- The safety and effectiveness have not been studied in the following patient populations:
 - Pregnant and nursing women
 - Individuals less than 18 years old
 - Individuals with feeding pedicle aneurysms not associated with the nidus, or distal feeders to the nidus or vascular malformation.
- Some data indicate that dimethyl sulfoxide potentiates other concomitantly administered medications.
- A garlic-like taste may be noted by the patient with use of the PHIL device due to the DMSO component. This taste may last several hours. An odor on the breath and skin may be present.
- Inspect product packaging prior to use. Do not use if sterile barrier is open or damaged.
- Use prior to expiration date.
- Verify that the microcatheters and accessories (see directions for use) used in direct contact with the PHIL polymer are clean and compatible with the material and do not trigger polymerization or degrade with contact. Use only DMSO compatible microcatheters indicated for use in the neuro or peripheral vasculature, and MicroVention sterile, prefilled syringes. Other microcatheters or syringes may not be compatible with DMSO and their use can result in thromboembolic events due to microcatheter degradation. Refer to the Warnings and Directions for Use sections.
- Upon completion of the PHIL device injection, wait three minutes, slightly aspirate the syringe, and then gently pull the microcatheter to separate it from the PHIL cast. Failure to wait this recommended time to retrieve the microcatheter after injection may result in fragmentation of the PHIL device into non-target vessels.

1.4.3 Potential Complications / Adverse Effects

The potential complications / adverse effects for the PHIL device include, but are not limited to:

- Hematoma at the puncture site
- Non-target arterial thrombosis
- Ischemic events due to embolic migration, vasospasm, thrombosis
- Hemorrhagic accidents: vascular rupture – perforation
- Hemodynamic changes induced by the embolization may result in hemorrhagic complications

- These ischemic or hemorrhagic complications may result in various functional neurological deficits, stroke, and possibly death.

1.4.4 Other Aspects of Safety

Data relevant to the clinical safety and performance of the PHIL devices was collected and evaluated from routine data sources from PMS such as complaints, Corrective and Preventative Action (CAPA) and SCAR (01 Jan 2020 to 31 Mar 2024), as well as post-market clinical follow-up (PMCF).

The data has demonstrated the clinical safety of the subject devices:

- There were 28,229 units shipped between 01 March 2020 to 29 February 2024, and there were 61 complaint records, for an overall complaint rate of 0.216%.
- There were zero (0) MDR events associated with the subject device during the PMS review period and four (4) MDV events.
- During the current review period three (3) SCARs were opened or in process that pertained to the PHIL devices. All SCARs have been completed with 100% effectiveness.
- During the current review period, twelve (12) CAPAs were opened or in process that pertained to the PHIL devices. As of this document's preparation, nine (9) CAPAs have been closed, two (2) are in Implementation, and one (1) is in Investigation.

The clinical evidence generated through this PMS will be used in the clinical evaluation of PHIL devices, with the aim of confirming the safety and performance throughout the expected lifetime of the device, of ensuring the continued acceptability of identified risks and of detecting emerging risks.

1.5 Summary of the Clinical Evaluation and PMCF

A Clinical Evaluation of the PHIL liquid embolic system is continuously updated and conducted in accordance with the requirements in MEDDEV.2.7.1 Revision 4– Guidelines on Medical Devices – Evaluation of Clinical Data: A Guide for Manufacturers and Notified Bodies. It includes the following:

- Literature Based Safety Appraisal
- A search of published relevant and available scientific literature was performed to assess the risks and benefits associated with other competitive predicate devices.
- Summary of Clinical Studies
- MicroVention has gathered data from post-market Trials/ Studies/ Registries under its sponsorship in which PHIL liquid embolic system were utilized
- Performance and Safety - Design Verification and Validation Data Analysis
- Product Literature and Instructions for Use

- The CER includes the methodology, literature references and conclusions and are reviewed and signed by an appropriately qualified physician.

The Clinical Evaluation Report, documents available clinical data relevant to the PHIL liquid embolic system. The available clinical data was collected, appraised, and analyzed, and it was determined that there is sufficient clinical evidence on the safety and performance in accordance with the intended purpose.

The Clinical Evaluation Report documents the benefit-risk profile, including side-effects, in the intended target patient populations and medical indications by assessing the clinical evidence against the hazards and patient harms as informed by the Risk Management and Post-Market Surveillance (PMS) documentation. The report also demonstrates the acceptability of the benefit-risk profile based on the current knowledge and state of the art in the concerned medical fields. Therefore, the clinical evaluation has established that the available clinical data are sufficient to establish conformity with all relevant Safety and Performance Requirements (Annex I) of EU MDR 2017/745 and confirm the safety and performance of the PHIL liquid embolic system.

In addition, Post-Market Surveillance (PMS) is a continuous process at MicroVention to gather, record, and analyze relevant data on the quality, performance and safety of a device throughout its entire lifetime actively and systematically. The planning and execution of PMS are conducted in accordance with the European Medical Device Regulation (MDR (EU) 2017/745), Chapter VII, Section 1 Post-Market Surveillance and MicroVention Post Market Procedures.

Given the evidence and data presented in the clinical evaluation and post market surveillance, and when the PHIL liquid embolic system is used according to the manufacturer's Instructions for Use, the risk to benefit profile is deemed acceptable.

1.5.1 Equivalent Device Clinical Data

Equivalency is not claimed in the clinical evaluation for the PHIL liquid embolic system.

1.5.2 Pre-CE-Mark Clinical Data

There were no EU pre-market clinical studies conducted for the PHIL Liquid Embolic System devices.

1.5.3 Clinical Data

Clinical data sources to evaluate the safety and performance of the PHIL Liquid Embolic System was collected from the following reputable data sources:

- Post-Market Clinical Studies
 - PHIL in the Treatment of Intracranial Dural AVF
 - PHIL Evaluation in the Endovascular Treatment of Intracranial Dural AVF

- Multicenter Single-Arm Study to Evaluate the Safety and Efficacy of Embolization of Dural Arteriovenous Fistulas (dAVF) Using the PHIL Liquid Embolic System
- PHIL Embolic System Pediatric IDE (Study of PHIL Embolic System in the Treatment of Intracranial Dural Arteriovenous Fistulas in the Pediatric Population (NCT03731000))
- Published Peer-reviewed Clinical Literature

The literature review identified clinical data involving total 406 patients harboring brain AVM or DAVF, and various peripheral vascular lesions, including peripheral AVM or AVF, visceral aneurysms, or pseudoaneurysms, or arterial lesions, and hypervascular tumor, as well as post-EVAR endoleak, in a total of 22 articles, of which 406 patients were treated with PHIL device. Of all 22 articles, complete (100%) technical success was achieved in 21 studies/articles, except in 1 study (94.7%). Immediate occlusion of 52-100% was reported in all 22 studies, Final angiographic follow up including post multi treatment sessions (ranging between 1 month and 28.4 months) results showed complete angiographic occlusion of 94.8% - 100% in all 22 studies with 20 studies reported 100% complete occlusion and the rest 2 studies reported 94.8% and 97.5%, respectively. Peri-procedural and post-procedural complications with 2 deaths were reported in the literature data, among all reported complications, zero were deemed related to PHIL embolic agent, and the 2 deaths determined due to worsening of baseline conditions. Adverse events were reported in literature reporting PHIL embolization with zero AEs attributable to the PHIL device. The literature data presented excellent performance data showing near complete/high rates of technical success, immediate angiographic occlusion, and occlusion during follow-up periods with no recurrence. Furthermore, low rates of morbidity and mortality confirmed safety of PHIL embolization, therefore, demonstrating effectiveness and safety of PHIL embolization in published literature data. There were no systematic misuses or off-label uses of the subject device identified in this report.

- Relevant data collected from PMCF activities in the PMCF Report described as routine data sources were integrated into the above data sections. These data sources include,
 - Scientific Literature
 - Registries
 - Sponsor-initiated post-market clinical studies
 - Investigator-initiated post-market clinical studies
 - There were no additional PMCF activities initiated to address specific findings of the previous clinical evaluation.

1.5.4 Clinical Performance and Safety

Adverse events reported in the clinical data (published literature) associated with the use of the PHIL Embolic System included intracranial hemorrhage (ICH) of 0.74% (3/406), SAH of 0.25%

(1/406), target vessel rupture of 0.25% (1/406), non-target vessel embolization of 0.74% (3/406), and draining vein thrombosis or venous sinus thrombosis of 1.23% (5/406). All-cause mortality up to 6 months was 0.49% (2/406) mainly due to worsening baseline condition.

The AEs associated with the use of the PHIL Embolic System reported in the published literature were generally categorized in correspondence with the risks identified during the risk assessment process. There were no new risks identified in the published literature that were not already considered through the risk management process.

The clinical benefits of the PHIL Embolic System in embolization of peripheral and neurovascular lesions including bAVM, dAVF, peripheral AVM, various peripheral vascular lesions, as well as hypervascular tumors have been demonstrated by favorable angiographic and clinical outcomes, including complete (100%) technical success achieved in 95.5% (21/22) of the studies, immediate occlusion of 52-100% reported in all 22 studies, angiographic follow up (ranging from 1 -28.4 months) after multi treatment sessions, of 94.8% - 100% complete occlusion in all 22 articles, and favorable clinical follow up (ranging from 1month to 28.4 months) in all applicable studies.

The clinical benefits are measured with technical success of the PHIL embolic agent implantation, immediate angiographic vascular lesion occlusion, as well as follow up angiographic occlusion and clinical outcomes evaluated with improvement of clinical symptoms, AVM/AVF downgrading, recanalization / recurrence, or neurological function mRS scores (0-2). The clinical literature data reported the PHIL Embolic System use was associated with high rates of technical success, immediate angiographic occlusion or nidus obliteration, favorable follow up angiographic occlusion, and zero recurrence. These results are in line with the results generated from alternative treatments or the state of the art.

Overall, treatment using the PHIL Embolic System has been associated with a high level of procedural safety and effectiveness, which has been consistently observed in all existing prospective trials and published literature. The good angiographic and clinical outcomes (indicated with complete angiographic occlusion of the embolized vascular lesion, or complete nidus obliteration of the embolized AVM, or favorable mRS 0-2) as well as the morbidity and mortality reported in most published data associated with the use of the PHIL Embolic System were in line with the rates associated with similar liquid embolic agents.

1.5.5 Post-Market Clinical Follow-up

From the evidence provided in this clinical evaluation, no PMCF studies are required for the PHIL Liquid Embolic System. The level of clinical evidence presented in this report is sufficient to support conformity to the relevant Essential Requirements, including a favorable benefit/risk ratio. No potential residual risks and/or unclarity on long term clinical performance that may impact the benefit/risk ratio were identified. No concerns were identified regarding the benefit-risk determination, the consistency of that evidence with the intended purpose, including the medical indication(s). This clinical evaluation has demonstrated that the PHIL Liquid Embolic System maintains an acceptable safety and performance profile and did not identify any questions relating

to clinical safety or performance (i.e., residual risks) when used in accordance with its approved labelling.

1.6 Possible Diagnostic or Therapeutic Alternatives

1.6.1 Treatment Options and Interventions

Endovascular embolization is a minimally invasive procedure that blocks blood vessels supplying a tumor or abnormal vessels. It is done while preserving normal blood flow in the surrounding regions to maintain tissue health while concurrently dealing with problematic bleeding.

Treatment with an embolization procedure results in the deliberate obstruction of a blood vessel for a medical benefit. Catheter-directed embolization of peripheral and neurovascular AVMs and other lesions, including AVFs and hypervascular tumors, has been shown to result in major clinical improvement or complete disappearance of symptoms in the vast majority of patients. During these procedures, the blood vessel of interest is obstructed with an embolic agent. The type of agent utilized depends on a number of factors, including the type and size of blood vessel, and the need for a temporary or permanent blockage of the blood vessel. Embolic materials include:

- Particulates
- Mechanical
- Liquids/Gels

Particulates

Particulates are the most commonly used embolic agent due to the variety of available sizes, shapes, and properties (permanent, biodegradable, natural, synthetic, etc.), with microspheres often preferred due to their controllable shape and size distribution (Hu et al., 2019, Santoro et al., 2019). Particulates are delivered proximal to the target vasculature via a catheter. The particles are then driven, via blood flow, into the smallest they can fit. Once lodged within the vessel, they mechanically obstruct flow and thrombus formation begins. Particulates are typically used when the target vasculature for thrombosis formation are small vessels including capillaries. Drawbacks include unintentional occlusion of larger proximal vessels due to clumping of particles and downstream embolization due to size and shape irregularity.

Mechanical

Mechanical embolization devices include coils and vascular plugs that allow complete occlusion of a specific vascular site. These devices are used in vessels of various sizes, including large vessels (Hu et al., 2019, Santoro et al., 2019). Coils and plugs are typically permeant implants with bioinert metallic cores (e.g., stainless steel, platinum, nitinol). During the procedure, the implant is placed in the desired location via a catheter. Once positioned, the implant expands into its preformed geometry in the vessel lumen. The expansion of the device creates radial force for vessel wall apposition, anchoring the device in place. The embolization of the vessel is achieved, in the use of coils by filling the space within the anchoring coil with additional coils, if needed. An occlusive membrane supported by the braided wire mesh core results in embolization when plugs are used.

The positioning of the coils or plug is then confirmed by radiography. Both coils and plugs are also used in conjunction in some cases (Adachi et al., 2020, Jardin et al., 2020).

Mechanical embolization devices are used to treat pulmonary and peripheral AVMs (Adachi et al., 2020, Bailey et al., 2019, Greene and Burrows, 2020, Gupta et al., 2020, Hawkins and Cheung, 2019), as well as a number of other issues.

Liquids/Gel

Liquid/gel embolic agents used in peripheral interventions have the ability to penetrate into smaller or more complex vascular targets where catheters and larger embolic therapeutics cannot (Hu et al., 2019, Santoro et al., 2019). This type of embolic agent blocks blood flow by chemically forming a permanent cast and does not depend on the patient's coagulation system for thrombosis formation. One type of liquid/gel embolics are sclerosing agents, ethanol being the most common, which functions by denaturing vessel endothelium and causing necrosis of vascular structures, such as AVMs (Kim et al., 2020, Bouwman et al., 2020). Other forms of liquid/gel embolic include in-situ polymerization (TRUFILL) which is commonly known as "glue" and *in-situ* precipitating fluids (Onyx, PHIL) which create casts when exposed to body fluids in the target vasculature. Drawbacks of liquid/gel embolics include cardiopulmonary complications (ethanol), necrosis of non-target tissue, nerve injury, distal thrombolysis, as well as with agents such as Onyx, high cost, radiopacity and residual mass effect (Bouwman et al., 2020, Kim et al., 2021).

Unlike other liquid embolics, the PHIL system radiopacifier is an integral part of the co-polymer (homogenous radiopacification); the inclusion of a nonmetallic radiopaque agent results in fewer artifacts with CT or magnetic resonance imaging (MRI) imaging.

Interventions for AVMs and Hypervascular Tumors

Treatment for AVMs and hypervascular tumors depends on a number of factors. Is the goal to cure the issue with an embolization technique, or reduce blood flow to make surgical removal safer? Current treatment options include the microsurgical removal of the AVM, or radiation therapy, which uses X-rays and other high-energy rays to destroy the cells at the AVM. In addition, both AVMs (and AVFs) and hypervascular tumors can be treated via embolization. Embolization blocks the blood flow in the vessels targeted, leading to a reduction in the AVM or ABF or reducing or stopping blood flow to the tumor (Pierot et al., 2013, Triano et al., 2020, Vollherbst et al., 2022).

Liquid embolic agents are used to induce occlusions in a region of a blood vessel of interest. These agents polymerize after contact with blood to form a solid material. The PHIL liquid embolic system is intended for use in the embolization of lesions in the peripheral and neurovasculature. Liquid embolics are utilized when other treatments, such as medications and life-style changes, are no longer a viable option (Lucatelli et al., 2021, Vollherbst et al., 2022).

Arteriovenous Malformations

Treatment for AVMs depends on where the abnormality is found, the patient’s signs and symptoms and overall health, and the risk of treatment. Sometimes, an AVM is monitored with regular imaging tests to watch for changes or problems. Other AVMs require treatment. Determining whether an AVM needs treatment involves factors including whether the AVM:

- Has bled
- Is causing symptoms other than bleeding
- Is in a part of the brain where treatment can be safely provided

The main treatment for AVM is surgery. The surgery might completely remove the AVM. This treatment is usually used when the AVM is in an area where surgeons can safely and easily remove the AVM, with little risk of causing significant damage to the brain tissues. As mentioned above, endovascular embolization, such as with a liquid embolic agent, is a type of surgery that is also utilized to treat AVMs (Giurazza et al., 2021). Stereotactic radiosurgery using highly focused beams of radiation to damage the blood vessels and stop the blood supply to the AVM, may also be used to treat AVMs (Chen et al., 2020).

Hypervascular Tumors

A hypervascular tumor is a tumor that has an abnormally large number of blood vessels attached. The elevated number of blood vessels may increase the risk of bleeding, and for this reason hypervascular tumors can often be difficult to remove. Reducing or eliminating this excessive blood flow feeding the tumor aids in the treatment. Liquid embolic agents are effectively utilized to treat these tumors, reducing or stopping the excessive blood flow to the tumor by blocking the vessels targeted (Pedicelli et al., 2020).

1.6.2 Available Technologies

Liquid embolic systems such as those listed within the PHIL liquid embolic system, are well established medical devices with numerous types and styles available from a variety of manufacturers. A few examples of the liquid embolic systems similar to the PHIL liquid embolic system are provided in table below.

Table 1-4: Similar Devices

Similar Device	Manufacturer	Intended use
TruFill® n-BCA	Johnson & Johnson	The TRUFILL® n-BCA Liquid Embolic System is used under fluoroscopic guidance to obstruct or reduce blood flow to cerebral arteriovenous malformations (AVMs) via super-selective catheter delivery when pre-surgical devascularization is desired
Onyx® LES	Medtronic	The Onyx® LES is used for presurgical embolization of brain arteriovenous malformations (bAVMs).

1.7 Suggested Profile and Training for Users

The PHIL Liquid Embolic System should be used only by physicians with neuro or peripheral interventional training and a thorough knowledge of the vascular pathology to be treated, vascular architecture, angiographic techniques, and super-selective embolization techniques.

1.8 Reference to any Harmonized Standards and CS

Common specification(s) to comply with:
Not Applicable
Harmonized standard(s) to comply with:

Standards	Edition	Standard Title
Quality System		
EN ISO 13485	2016/A11:2021	Medical devices - Quality management systems - Requirements for regulatory purposes (ISO 13485:2016)
Risk Management		
EN ISO 14971	2019/A11:2021	Medical devices - Application of risk management to medical devices (ISO 14971:2019)
EN IEC 60812	2018	Failure modes and effects analysis (FMEA and FMECA) (IEC 60812:2018)
Usability		
EN ISO 62366-1	2015/A1: 2020	Medical devices – Part 1: Application of usability engineering to medical devices (IEC 62366- 1:2015/A1:2020)
Clinical		
EN ISO 14155	2020	Clinical investigation of medical devices for human subjects -- Good clinical practice (ISO 14155:2020)
Post Market Surveillance		
ISO/TR 20416	2020	Medical Devices-Post Market Surveillance for Manufacturers
Labeling		
EN ISO 15223-1	2021	Medical devices - Symbols to be used with information to be supplied - Part 1: General requirements (ISO 15223-

Standards	Edition	Standard Title
		1:2021)
EN ISO 20417	2021	Medical devices - Information supplied by the manufacturer of medical devices (ISO 20417:2021, Corrected version 2021-12)
Packaging		
EN ISO 11607-1	2020/A11:2022	Packaging for Terminally Sterilized Medical Devices Part 1: Requirements for Material, Sterile Barrier Systems and Packaging Systems (ISO 11607-1:2019)
EN ISO 11607-2	2020/A11:2022	Packaging for terminally sterilized medical devices - Part 2: Validation requirements for forming, sealing and assembly processes (ISO 11607-2:2019)
ISTA 3A	2018	Packaged-Products for Parcel Delivery System Shipment 150 lbs. (70 kg) or Less
ASTM D4169	2023e1	Standard Practice for Performance Testing of Shipping Containers and Systems
ASTM F88	2023	Standard Test Method for Seal Strength of Flexible Barrier Materials
ASTM F1886	2016	Standard Test Method for Determining Integrity of Seals for Flexible Packaging by Visual Inspection
ASTM F1929	2023	Standard Test Method for Detecting Seal Leaks in Porous Medical Packaging by Dye Penetration
ASTM F2096	2011/R2019	Standard Test Method for Detecting Gross Leaks in Packaging by Internal Pressurization (Bubble Test)
Shelf Life & Stability		
ASTM F1980	2016	Standard Guide for Accelerated Aging of Sterile Barrier Systems for Medical Devices
Biocompatibility		
EN ISO 10993-1	2020	Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process (ISO 10993-1:2018, including corrected version 2018-10)
EN ISO 10993-3	2014	Biological evaluation of medical devices - Part 3: Tests for genotoxicity, carcinogenicity and reproductive toxicity (ISO 10993-3:2014)

Standards	Edition	Standard Title
EN ISO 10993-4	2017	Biological evaluation of medical devices - Part 4: Selection of tests for interactions with blood (ISO 10993-4:2017)
EN ISO 10993-5	2009	Biological evaluation of medical devices - Part 5: Tests for in vitro cytotoxicity (ISO 10993-5:2009)
EN ISO 10993-6	2016	Biological evaluation of medical devices - Part 6: Tests for local effects after implantation (ISO 10993-6:2016)
EN ISO 10993-10	2023	Biological evaluation of medical devices - Part 10: Tests for skin sensitization (ISO 10993-10:2021)
EN ISO 10993-11	2018	Biological evaluation of medical devices - Part 11: Tests for systemic toxicity (ISO 10993-11:2017)
EN ISO 10993-12	2021	Biological evaluation of medical devices - Part 12: Sample preparation and reference materials (ISO 10993-12:2021)
EN ISO 10993-18	2020/A1:2023	Biological evaluation of medical devices - Part 18: Chemical characterization of medical device materials within a risk management process (ISO 10993-18:2020/Amd 1:2022)
EN ISO 10993-23	2021	Biological evaluation of medical devices - Part 23: Tests for irritation (ISO 10993-23:2021)
Manufacturing (Environmental Controls)		
EN ISO 14644-1	2015	Cleanrooms and associated controlled environments - Part 1: Classification of air cleanliness by particle concentration (ISO 14644-1:2015)
EN ISO 14644-2	2015	Cleanrooms and associated controlled environments - Part 2: Monitoring to provide evidence of cleanroom performance related to air cleanliness by particle concentration (ISO 14644-2:2015)
ANSI/AAMI ST72	2019	Bacterial endotoxins – Test methods, routine monitoring, and alternatives to batch testing
Sterilization		
EN 556-1	2001/AC:2006	Sterilization of medical devices – Requirements for medical devices to be designated ‘STERILE’ – Part 1: Requirements for terminally sterilized medical devices
EN ISO 11737-1	2018/A1:2021	Sterilization of health care products - Microbiological methods - Part 1: Determination of a population of microorganisms on products (ISO 11737-1:2018/Amd 1:2021)

Standards	Edition	Standard Title
EN ISO 11737-2	2020	Sterilization of health care products - Microbiological methods - Part 2: Tests of sterility performed in the definition, validation and maintenance of a sterilization process (ISO 11737-2:2019)
Biological Indicators		
EN ISO 11138-1	2017	Sterilization of health care products - Biological indicators - Part 1: General requirements (ISO 11138-1:2017)
Steam/Autoclave/Moist Heat		
EN ISO 17665-1	2006	Sterilization of health care products — Moist heat — Part 1: Requirements for the development, validation and routine control of a sterilization process for medical devices (ISO 17665- 1:2006)
Device Specific		
Implants		
EN ISO 14630	2012	Non-active surgical implants - General requirements (ISO 14630:2012)
EN ISO 25539-1	2017	Cardiovascular implants - Endovascular devices- Part 1: Endovascular prostheses (ISO 25539- 1:2017)
ISO 16428	2005	Implants for surgery – Test solutions and environmental conditions for static and dynamic corrosion tests on implantable materials and medical devices
ASTM F2129	2019a	Standard Test Method for Conducting Cyclic Potentiodynamic Polarization Measurements to Determine the Corrosion Susceptibility of Small Implant Devices
ASTM F3044	2020	Standard Test Method for Evaluating the Potential for Galvanic Corrosion for Medical Implants
ASTM G16	2013	Standard Guide for Applying Statistics to Analysis of Corrosion Data
Radiopacity		
ASTM F640	2023	Standard Test Methods for Determining Radiopacity for Medical Use

1.9 References

- ADACHI, A., OHTA, K., JAHANGIRI, Y., MATSUI, Y., HORIKAWA, M., GEERATIKUN, Y., CHANSANTI, O., YATA, S., FUJII, S. & STEINBERGER, J. 2020. Treatment of pulmonary arteriovenous malformations: Clinical experience using different embolization strategies. *Japanese journal of radiology*, 1-5.
- BAILEY, C. R., ARUN, A., TOWSLEY, M., CHOI, W. K., BETZ, J. F., MACKENZIE, S., ABOU AREDA, M., DUVVURI, M., MITCHELL, S. & WEISS, C. R. 2019. MVP™ micro vascular plug systems for the treatment of pulmonary arteriovenous malformations. *Cardiovascular and interventional radiology*, 42, 389-395.
- BOUWMAN, F. C., BOTDEN, S. M., VERHOEVEN, B. H., KOOL, L. J. S., VAN DER VLEUTEN, C. J., DE BLAAUW, I. & KLEIN, W. M. 2020. Treatment Outcomes of Embolization for Peripheral Arteriovenous Malformations. *Journal of Vascular and Interventional Radiology*, 31, 1801-1809.
- CHEN, C. J., DING, D., DERDEYN, C. P., LANZINO, G., FRIEDLANDER, R. M., SOUTHERLAND, A. M., LAWTON, M. T. & SHEEHAN, J. P. 2020. Brain arteriovenous malformations: A review of natural history, pathobiology, and interventions. *Neurology*, 95, 917-927.
- GIURAZZA, F., CORVINO, F., SILVESTRE, M., CANGIANO, G., DE MAGISTRIS, G., CAVAGLIÀ, E., AMODIO, F. & NIOLA, R. 2021. Embolization of peripheral arteriovenous malformations and fistulas with precipitating hydrophobic injectable liquid (PHIL®). *Radiol Med*, 126, 474-483.
- GREENE, A. K. & BURROWS, P. E. 2020. Arteriovenous malformation. *Vascular Anomalies*. Springer.
- GUPTA, M., AGRAWAL, A., IAKOVOU, A., COHEN, S., SHAH, R. & TALWAR, A. 2020. Pulmonary artery aneurysm: a review. *Pulmonary circulation*, 10, 2045894020908780.
- HAWKINS, C. M. & CHEWNING, R. H. 2019. Diagnosis and Management of Extracranial Vascular Malformations in Children: Arteriovenous Malformations, Venous Malformations, and Lymphatic Malformations. *Semin Roentgenol*, 54, 337-348.
- HU, J., ALBADAWI, H., CHONG, B. W., DEIPOLYI, A. R., SHETH, R. A., KHADEMOSSEINI, A. & OKLU, R. 2019. Advances in biomaterials and technologies for vascular embolization. *Advanced Materials*, 31, 1901071.
- JARDINET, T., BONNE, L., OYEN, R. & MALEUX, G. 2020. Initial Experience With the Microvascular Plug in Selective Renal Artery Embolization. *Vascular and endovascular surgery*, 54, 240-246.
- KIM, R., DO, Y. S. & PARK, K. B. 2020. How to Treat Peripheral Arteriovenous Malformations. *Korean Journal of Radiology*, 21.
- KIM, R., DO, Y. S., PARK, K. B., KIM, D. I., HEO, S. H., PARK, H. S., SHIN, S. W., CHO, S. K. & HYUN, D. H. 2021. Therapeutic outcomes of embolotherapy of extremity bone intraosseous arteriovenous malformation with ethanol, coils, and n-butyl cyanoacrylate. *J Vasc Surg*, 73, 2090-2097 e5.

- LUCATELLI, P., CORONA, M., TEODOLI, L., NARDIS, P., CANNAVALE, A., ROCCO, B., TROBIANI, C., CIPOLLARI, S., ZILAHY DE GYURGYOKAI, S., BEZZI, M. & CATALANO, C. 2021. Use of Phil Embolic Agent for Bleeding in Non-Neurological Interventions. *J Clin Med*, 10.
- PEDICELLI, A., LOZUPONE, E., VALENTE, I., SNIDER, F., RIGANTE, M., D'ARGENTO, F., ALEXANDRE, A., GARIGNANO, G., CHIUMARULO, L., PALUDETTI, G. & COLOSIMO, C. 2020. Pre-operative direct puncture embolization of head and neck hypervascular tumors using SQUID 12. *Interv Neuroradiol*, 26, 346-353.
- PIEROT, L., COGNARD, C., HERBRETEAU, D., FRANSEN, H., VAN ROOIJ, W. J., BOCCARDI, E., BELTRAMELLO, A., SOUROUR, N., KUPCS, K., BIONDI, A., BONAFÉ, A., REITH, W. & CASASCO, A. 2013. Endovascular treatment of brain arteriovenous malformations using a liquid embolic agent: results of a prospective, multicentre study (BRAVO). *Eur Radiol*, 23, 2838-45.
- SANTORO, J., DI MARTINO, M., PIOTTO, S. & MIDULLA, M. Image-Guided Mini-Invasive Treatments for Vascular and Oncologic Diseases: Embolization Therapy. International Conference on Bio and Nanomaterials, 2019. Springer, 233-242.
- TRIANO, M. J., LARA-REYNA, J., SCHUPPER, A. J. & YAEGER, K. A. 2020. Embolic Agents and Microcatheters for Endovascular Treatment of Cerebral Arteriovenous Malformations. *World Neurosurg*, 141, 383-388.
- VOLLHERBST, D. F., CHAPOT, R., BENDSZUS, M. & MÖHLENBRUCH, M. A. 2022. Glue, Onyx, Squid or PHIL? Liquid Embolic Agents for the Embolization of Cerebral Arteriovenous Malformations and Dural Arteriovenous Fistulas. *Clin Neuroradiol*, 32, 25-38.