Teladoc Health™ Lite[®] 4 Quick Start Guide

P/N: MA-20145-002 © Copyright, 2021 Teladoc Health™ Release: 042



Table of Contents

LITE 4 PRE-INSTALLATION	3
UNPACKING AND CHARGING THE LITE 4	.4
LITE 4 WI-FI CONNECTION SETUP	5
VERIFY DEVICE CONNECTIVITY WITH TELADOC HEALTH NETWORK	. 5
CONTACT INFORMATION	8



Lite 4 Pre-Installation

NOTES:

- The Lite 4 should be installed by an IT professional to ensure proper setup and network configuration.
- Teladoc Health Manuals and documents are available at: <u>https://intouchhealth.com/support/device-resources/</u>.

Review the following documents for network requirements:

- MB-14011 Teladoc Health Telehealth Network Specifications
- MB-15513 Network Configuration for Teladoc Health Devices

Plan the installation of the Lite 4, making sure the following are available during the installation:

CAUTION: Guest, Staff, and VIP type networks are not suitable for connection of a medical device peripheral.

- The clinical, secure WiFi network name and credentials.
- All locations where the device is to be used.
- Any peripherals to be used with the Lite 4 such as a stethoscope.

NOTE: Allow the Lite 4 to fully charge before starting the installation.



Unpacking and Charging the Lite 4

WARNING: Plug the Lite 4 into a grounded "Hospital Grade" electrical outlet to minimize the risk of electrical shock.

- 1. Carefully remove the Lite 4 from its packaging, take care not to cause damage.
- 2. Flip the battery switch to the right.
 - Pressing the button to the left will disconnect the batteries, and pressing to the right will connect the batteries.



- 3. Plug the Lite 4 into grounded "Hospital Grade" AC outlet and allow the system to charge for at least 6 hours to reach a full charge.
- 4. Power on the Lite 4 by pressing the power button located on the back of the head.






Lite 4 Wi-Fi Connection Setup



2. Select the desired secure, clinical wireless SSID network.



(missing or bad snippet)

3. Enter the network security key.

						E	nter pa	asswo	rd					
		2	3	4	5	6	ī	7	8	9	0			×
ta	ab	q	w	е			}	'				р	ſ] /
са	ips	а	s	d		g	ł	ı		k				8
	shift		z	x		v	b		n	n				shift
cle	ear						spa	ace						clear

4. Use the **Information Icon** (I) on the lower right corner of the screen and ensure that the status is **Ready** before attempting to make a connection.

NOTES:

- The Diagnostic Information page will display current WiFi status.
- The Lite 4 will automatically connect to known Wi-Fi profiles. For stored WiFi profiles, the password can be changed by selecting it and editing the password.

Verify Device Connectivity with Teladoc Health Network


NOTE: These steps must be completed prior to granting connectivity between the Provider Access Software and the Lite 4 for clinical use.



1. Open the Lite 4 information screen by selecting the 🛈 icon from the idle screen.

nTouch Lite		Vetwork details	
Ready Battery charge	100%	Connection type	₩iFi
	ROBOT-19024	WiFi SSID	TIC
	Demo-Lite 4 (ED 1)	Signal strength	88%
	Lite V4	External IP	64.29.227.1
	10.40.11	Internal IP	192.168.12.69
	January 03 2020		
	Littmann		
	Thermal		
Connectivity sta	tus		
	Registered		
FMS	Registered		
	Registered		
	Registered		
	Trogiscorea		

- 2. Contact your Teladoc Health account manager to complete the installation of the Lite 4 and to allow it to be available for use.
- 3. Select the 🖸 icon from the bottom of the Diagnostic Information Screen.
- 4. Select the icon to open and run the Network Checkup.
 - Successful test results will display a green check mark \searrow
 - Unsuccessful tests results will display a red 🎴





- 5. Follow the instructions on the screen should there be any unsuccessful test results.
- 6. Select the Contorer and Technical Services at +1 (877) 484-9119.



Contact Information

24/7 Live Technical Support

<u>+1 (877) 484-9119</u>

24/7 Live Remote Technical Support & Live Chat

https://intouchhealth.com/contact-us/

Email Support

ITHSupport@intouchhealth.com

Teladoc Health User Manuals

https://intouchhealth.com/manuals/

Sales & Product Demos

(805) 562-8686



InTouch Technologies, Inc. 7402 Hollister Avenue Goleta, CA 93117 Ph: +1.805.562.8686 • Fax: +1.805.562.8663

Teladoc

ΗΓΔΙΤΗ

©2021 Teladoc Health, Inc. InTouch Health and InTouch Technologies are now a Teladoc Health company, and InTouch Health is a registered trademark of Teladoc Health.

LEARN MORE

TeladocHealth.com | engage@teladochealth.com

About Teladoc Health

Teladoc Health is the global virtual care leader, helping millions of people resolve their healthcare needs with confidence. Together with our clients and partners, we are continually modernizing the healthcare experience and making high-quality healthcare a reality for more people and organizations around the world.