

# Approved Manufacturers

Clinell Peracetic Acid Wipes\* have been approved for use by the following manufacturers:

BRAND	EQUIPMENT
<b>Abecca</b>	Positioning Supports & Mattresses
<b>Active Key GmbH &amp; Co. KG</b>	All Active Key MedicalKey™ products, including: AK-C4112, AK-C4412, AK-C7012 & AK-C7412
<b>AGFA</b>	DX-D products
<b>Agile Medical</b>	All products
<b>BD (Bard)</b>	Site-Rite® 8 Ultrasound System and Probe
<b>Carrflex</b>	Mattress Covers, Upholstery Tops & Associated Equipment
<b>Carr Plastics</b>	Techmaflex
<b>Dartex Coating</b>	Mattresses (with polyurethane face)
<b>Fresenius Kabi - Agilia range</b>	VP MC volumetric infusion pump SP MC syringe infusion pump SP TIVA TCI syringe infusion pump SP PCA syringe infusion pump
<b>Fujifilm Sonosite</b>	Ultrasound Equipment (Edge, NanoMaxx, MicroMaxx, TITAN) iViz devices and C60v, L25v, L38v & P21v probe
<b>Howard Wright</b>	Mattresses

\*Previously Clinell Sporicidal Wipes.

# Approved Manufacturers

Clinell Peracetic Acid Wipes\* have been approved for use by the following manufacturers:

BRAND	EQUIPMENT
<b>HP</b>	HP EliteBook 840 G6 Healthcare Edition HP EliteOne 800 G5 Healthcare Edition All-in-One HP Healthcare Edition HC270cr Clinical Review Display HP Healthcare Edition HC241 Clinical Review Monitor HP Healthcare Edition HC241p Clinical Review Monitor HP Healthcare Edition HC271 Clinical Review Monitor HP Healthcare Edition HC271p Clinical Review Monitor
<b>Masimo</b>	Root, Radical-7, Rad-5/Rad-57/Pronto, Ra-8, SedLine Module, O3 Module Rad-67, Rad-97 (remaining residue must be wiped off) Cable (PVC jacket), Cable (TPU jacket)
<b>Mindray</b>	N Series Monitor
<b>Modsel</b>	All product ranges
<b>OES Medical</b>	Pressure Monitor and UAM Anaesthetic Machine
<b>Orfit Industries</b>	All related types of AIO 3.0 Low Density Head Supports Medium Density Blue Head Support
<b>Panaz</b>	Cadet Furniture
<b>Parity Medical</b>	Mobile Computer Carts
<b>Recticel</b>	Mattresses

\*Previously Clinell Sporidical Wipes.

# Approved Manufacturers

Clinell Peracetic Acid Wipes\* have been approved for use by the following manufacturers:

BRAND	EQUIPMENT
<b>Philips</b>	Ultrasound Equipment Intellivue Monitor (remaining residue must be wiped off) Ultrasound Transducers (Please email <a href="mailto:info@gamahealthcare.com">info@gamahealthcare.com</a> for full list of compatible transducers)
<b>Siemens</b>	Siemens RapidPoint Systems Mammomat Inspiration
<b>Sidhil</b>	IQ Bed Frame
<b>Recticel</b>	Mattresses
<b>Siemens</b>	Siemens RapidPoint Systems Mammomat Inspiration
<b>Sidhil</b>	IQ Bed Frame
<b>Toshiba</b>	Ultrasound Equipment
<b>Tough-PAC</b>	iPad Covers
<b>Universal Ferrule Remover</b>	Ferrule Remover
<b>Zoll</b>	AED

\*Previously Clinell Sporocidal Wipes.